

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD010920\
 Data File : PD056893.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jan 2020 19:10
 Operator : AJ\MA
 Sample : INDA401
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA401

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 10 06:04:57 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD010920CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Jan 10 06:00:03 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.285	3.959	31618428	58312220	38.797	39.244
27)	SA Decachlor...	7.953	8.992	72406096	111.3E6	76.179	77.188
Target Compounds							
2)	A alpha-BHC	3.711	4.346	47455522	85873238	40.791	41.089
3)	MA gamma-BHC...	3.986	4.628	45201089	81947984	39.485	40.562
4)	MA Heptachlor	4.271	5.134	42556013	82904848	39.595	39.692
9)	A Endosulfan I	5.276	6.176	36421887	68606577	39.654	39.418
13)	MA Dieldrin	5.515	6.431	83781827	153.1E6	81.150	80.844
14)	MA Endrin	5.768	6.648	68926294	123.0E6	79.039	79.758
16)	A 4,4'-DDD	5.902	6.770	62310803	116.1E6	80.793	81.187
17)	MA 4,4'-DDT	6.138	7.068	62342335	118.1E6	85.176	82.848
20)	A Methoxychlor	6.686	7.525	155.1E6	307.0E6	388.475	392.899

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD010920\
Data File : PD056893.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jan 2020 19:10
Operator : AJ\MA
Sample : INDA401
Misc :
ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 10 06:04:57 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD010920CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jan 10 06:00:03 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

