

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD011019\
 Data File : PD050962.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 09 Jan 2019 19:56
 Operator : AJ\SJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 10 03:57:46 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD011019.M
 Quant Title : GC Extractables
 QLast Update : Thu Jan 10 03:41:46 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.227	4.047	31541390	623.0E6	21.117	20.936
28)	SA Decachlor...	7.834	9.164	23614747	316.3E6	19.456	15.854
Target Compounds							
3)	MA gamma-BHC...	3.927	0.000	1626301	0	0.690	N.D. #
6)	B beta-BHC	4.101f	4.939f	630110	29113051	0.665	1.609 #
8)	B Heptachlo...	0.000	5.930	0	75507205	N.D.	2.166 #
19)	B Endosulfa...	6.241f	0.000	816505	0	0.417	N.D. #
23)	Chlordane-1	4.081	0.000	504192	0	7.386	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD011019\
Data File : PD050962.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 09 Jan 2019 19:56
Operator : AJ\SJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 10 03:57:46 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD011019.M
Quant Title : GC Extractables
QLast Update : Thu Jan 10 03:41:46 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.227 min
Delta R.T.: 0.000 min
Response: 31541390
Conc: 21.12 ng/ml

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.047 min
Delta R.T.: 0.000 min
Response: 623048678
Conc: 20.94 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 3.927 min
Delta R.T.: 0.011 min
Response: 1626301
Conc: 0.69 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 0.000 min
Exp R.T. : 4.727 min
Response: 0
Conc: N.D.



#6 beta-BHC


R.T.: 4.101 min
Delta R.T.: -0.061 min
Response: 630110
Conc: 0.67 ng/ml

Instrument :
ECD_D
ClientSampleId :




#6 beta-BHC

R.T.: 4.939 min
Delta R.T.: 0.043 min
Response: 29113051
Conc: 1.61 ng/ml




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T. : 4.855 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 5.930 min
Delta R.T.: -0.006 min
Response: 75507205
Conc: 2.17 ng/ml



#19 Endosulfan Sulfate


R.T.: 6.241 min
Delta R.T.: -0.064 min
Response: 816505
Conc: 0.42 ng/ml

Instrument :
ECD_D
ClientSampleId :




#19 Endosulfan Sulfate

R.T.: 0.000 min
Exp R.T. : 7.330 min
Response: 0
Conc: N.D.




#23 Chlordane-1

R.T.: 4.081 min
Delta R.T.: 0.030 min
Response: 504192
Conc: 7.39 ng/ml



#23 Chlordane-1


R.T.: 0.000 min
Exp R.T. : 5.051 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 7.834 min
Delta R.T.: 0.000 min
Response: 23614747
Conc: 19.46 ng/ml

Instrument :
ECD_D
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 9.164 min
Delta R.T.: 0.000 min
Response: 316348514
Conc: 15.85 ng/ml