

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD011124CLP-TCLP.M
 Title : GC Extractables
 Last Update : Thu Jan 11 00:23:45 2024
 Response Via : Initial Calibration

Calibration Files

3 =PD080753.D 1 =PD080751.D 2 =PD080752.D
 4 =PD080754.D 5 =PD080755.D 6 =PD080756.D

Compound	3	1	2	4	5	6	Avg	%RSD
1) SA Tetrachloro-m...	1.389	1.420	1.423	1.385	1.346	1.396	1.393 E6	2.01
2) A alpha-BHC							0.000	-1.00
3) MA gamma-BHC (Li...							0.000	-1.00
4) MA Heptachlor							0.000	-1.00
5) MB Aldrin							0.000	-1.00
6) B beta-BHC							0.000	-1.00
7) B delta-BHC							0.000	-1.00
8) B Heptachlor ep...							0.000	-1.00
9) A Endosulfan I							0.000	-1.00
10) B trans-Chlordane							0.000	-1.00
11) B cis-Chlordane							0.000	-1.00
12) B 4,4'-DDE							0.000	-1.00
13) MA Dieldrin							0.000	-1.00
14) MA Endrin							0.000	-1.00
15) B Endosulfan II							0.000	-1.00
16) A 4,4'-DDD							0.000	-1.00
17) MA 4,4'-DDT							0.000	-1.00
18) B Endrin aldehyde							0.000	-1.00
19) B Endosulfan Su...							0.000	-1.00
20) A Methoxychlor							0.000	-1.00
21) B Endrin ketone							0.000	-1.00
22) Toxaphene-1	1.441	1.555	1.505	1.442	1.563	1.504	1.502 E4	3.51
23) Toxaphene-2	1.837	1.858	1.851	1.806	1.994	1.962	1.885 E4	3.99
24) Toxaphene-3	5.152	5.036	5.070	5.232	5.514	5.788	5.299 E4	5.55
25) Toxaphene-4	3.766	3.779	3.801	3.915	4.174	4.369	3.967 E4	6.28
26) Toxaphene-5	2.884	2.804	2.740	2.961	3.206	3.316	2.985 E4	7.67
27) SA Decachlorobip...	1.172	1.234	1.226	1.162	1.176	1.126	1.182 E6	3.43

Signal #2 Calibration Files

3 =PD080753.D 1 =PD080751.D 2 =PD080752.D
 4 =PD080754.D 5 =PD080755.D 6 =PD080756.D

Compound	3	1	2	4	5	6	Avg	%RSD
1) SA Tetrachloro-m...	2.366	2.378	2.441	2.368	2.393	2.392	2.390 E6	1.16
2) A alpha-BHC							0.000	-1.00
3) MA gamma-BHC (Li...							0.000	-1.00
4) MA Heptachlor							0.000	-1.00
5) MB Aldrin							0.000	-1.00
6) B beta-BHC							0.000	-1.00
7) B delta-BHC							0.000	-1.00
8) B Heptachlor ep...							0.000	-1.00
9) A Endosulfan I							0.000	-1.00
10) B trans-Chlordane							0.000	-1.00
11) B cis-Chlordane							0.000	-1.00
12) B 4,4'-DDE							0.000	-1.00
13) MA Dieldrin							0.000	-1.00
14) MA Endrin							0.000	-1.00
15) B Endosulfan II							0.000	-1.00
16) A 4,4'-DDD							0.000	-1.00
17) MA 4,4'-DDT							0.000	-1.00
18) B Endrin aldehyde							0.000	-1.00
19) B Endosulfan Su...							0.000	-1.00
20) A Methoxychlor							0.000	-1.00
21) B Endrin ketone							0.000	-1.00
22) Toxaphene-1	1.762	1.718	1.869	1.724	1.752	1.765	1.765 E4	3.08

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD011124CLP-TCLP.M
 Title : GC Extractables
 Last Update : Thu Jan 11 00:23:45 2024
 Response Via : Initial Calibration

Calibration Files

3	=PD080753.D	1	=PD080751.D	2	=PD080752.D
4	=PD080754.D	5	=PD080755.D	6	=PD080756.D

	Compound	3	1	2	4	5	6	Avg		%RSD
23)	Toxaphene-2	1.839	1.879	1.891	1.850	1.835	1.853	1.858	E4	1.20
24)	Toxaphene-3	5.891	5.218	5.750	6.210	6.554	6.898	6.087	E4	9.85
25)	Toxaphene-4	8.327	7.739	8.109	8.608	8.817	8.802	8.400	E4	5.07
26)	Toxaphene-5	3.300	3.108	3.293	3.544	3.600	3.616	3.410	E4	6.05
27) SA	Decachlorobip...	1.508	1.594	1.580	1.472	1.590	1.468	1.535	E6	3.87

(#) = Out of Range