

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD011124\
 Data File : PD080749.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jan 2024 14:55
 Operator : AR\AJ
 Sample : RESC018
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 RESC020

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 11 05:17:59 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD011124CLP-TCLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Jan 11 04:50:32 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.574	2.794	27996438	48488500	19.640	19.611
27)	SA Decachlor...	9.129	7.941	80012416	108.5E6	40.823	40.126
Target Compounds							
2)	A alpha-BHC	4.031	3.297	22713874	39211960	9.308	9.531
3)	MA gamma-BHC...	4.366	3.627	22054907	37275776	9.487	9.750
4)	MA Heptachlor	4.959	3.969	24236353	37205939	9.991	10.099
5)	MB Aldrin	5.302	4.250	23122350	37966450	10.070	10.422
6)	B beta-BHC	4.560	3.923	10441287	17360260	10.282	10.458
7)	B delta-BHC	4.810	4.154	23715361	38859320	10.122	9.809
8)	B Heptachlo...	5.730	4.752	21325411	34449866	9.698	10.091
9)	A Endosulfan I	6.116	5.123	21257649	30898315	10.023	9.598
10)	B trans-Chl...	5.986	5.003	21005035	32858623	9.836	9.921
11)	B cis-Chlor...	6.066	5.067	21659926	32543841	9.889	9.563
12)	B 4,4'-DDE	6.237	5.256	39075529	62200782	19.036	19.734
13)	MA Dieldrin	6.393	5.388	43857602	69451141	19.399	19.514
14)	MA Endrin	6.623	5.663	34598688	59895179	19.211	20.341
15)	B Endosulfa...	6.841	5.956	36958492	57299425	19.490	20.045
16)	A 4,4'-DDD	6.754	5.810	31534908	53538962	18.935	19.506
17)	MA 4,4'-DDT	7.070	6.061	33474716	53543195	19.540	19.602
18)	B Endrin al...	6.972	6.135	30157908	47688877	21.095	21.358
19)	B Endosulfa...	7.206	6.359	37233473	57305241	20.810	20.803
20)	A Methoxychlor	7.545	6.637	94890655	151.9E6	103.779	110.207
21)	B Endrin ke...	7.693	6.866	42374175	67853521	21.640	21.741

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD011124\
Data File : PD080749.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jan 2024 14:55
Operator : AR\AJ
Sample : RESC018
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 11 05:17:59 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD011124CLP-TCLP.M
Quant Title : GC Extractables
QLast Update : Thu Jan 11 04:50:32 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

