

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011625\
 Data File : PP069131.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Jan 2025 12:12
 Operator : YP\AJ
 Sample : Q1091-04 10X
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 FDA569G-1-2

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 16 13:06:21 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010625.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 07 11:05:55 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.669	3.980	1871624	1785616	1.345	1.829 #
2)	SA Decachlor...	10.517	9.103	2067770	2121860	1.852	1.766
Target Compounds							
3)	L1 AR-1016-1	0.000	5.117	0	125528	N.D.	3.710 #
4)	L1 AR-1016-2	0.000	5.117	0	125528	N.D.	2.494 #
7)	L1 AR-1016-5	0.000	5.597f	0	143275	N.D.	4.745 #
8)	L2 AR-1221-1	0.000	4.168	0	96113	N.D.	7.124 #
9)	L2 AR-1221-2	4.980	4.228f	122792	183332	9.540	17.665 #
10)	L2 AR-1221-3	4.980f	0.000	122792	0	3.167	N.D. #
11)	L3 AR-1232-1	4.980f	0.000	122792	0	4.137	N.D. #
12)	L3 AR-1232-2	0.000	5.117	0	125528	N.D.	5.248 #
15)	L3 AR-1232-5	0.000	5.597f	0	143275	N.D.	11.162 #
16)	L4 AR-1242-1	0.000	5.117	0	125528	N.D.	4.277 #
17)	L4 AR-1242-2	0.000	5.117	0	125528	N.D.	2.981 #
21)	L5 AR-1248-1	0.000	5.117	0	125528	N.D.	5.458 #
24)	L5 AR-1248-4	0.000	5.597f	0	143275	N.D.	3.481 #
29)	L6 AR-1254-4	0.000	6.624f	0	35237	N.D.	0.783 #
30)	L6 AR-1254-5	0.000	7.093	0	154628	N.D.	2.072 #
31)	L7 AR-1260-1	0.000	6.624f	0	35237	N.D.	0.615 #
32)	L7 AR-1260-2	0.000	6.754	0	154832	N.D.	2.350 #
34)	L7 AR-1260-4	0.000	7.442f	0	78347	N.D.	1.399 #
35)	L7 AR-1260-5	0.000	7.597f	0	64111	N.D.	0.522 #
36)	L8 AR-1262-1	0.000	7.182f	0	213479	N.D.	2.673 #
37)	L8 AR-1262-2	0.000	7.442f	0	78347	N.D.	1.120 #
38)	L8 AR-1262-3	0.000	7.893f	0	277459	N.D.	4.818 #
40)	L8 AR-1262-5	0.000	8.554f	0	88673	N.D.	1.816 #
41)	L9 AR-1268-1	0.000	7.893	0	277459	N.D.	1.710 #
43)	L9 AR-1268-3	0.000	8.231	0	149415	N.D.	1.125 #
44)	L9 AR-1268-4	0.000	8.554f	0	88673	N.D.	1.665 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011625\
Data File : PP069131.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Jan 2025 12:12
Operator : YP\AJ
Sample : Q1091-04 10X
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
FDA569G-1-2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 16 13:06:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010625.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 07 11:05:55 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.669 min
Delta R.T.: -0.003 min
Response: 1871624
Conc: 1.35 ng/ml

Instrument :
ECD_D
ClientSampleId :
FDA569G-1-2




#1 Tetrachloro-m-xylene

R.T.: 3.980 min
Delta R.T.: -0.003 min
Response: 1785616
Conc: 1.83 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.517 min
Delta R.T.: -0.008 min
Response: 2067770
Conc: 1.85 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.103 min
Delta R.T.: -0.008 min
Response: 2121860
Conc: 1.77 ng/ml



#3 AR-1016-1


R.T.: 0.000 min
Exp R.T.: 5.832 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
FDA569G-1-2




#3 AR-1016-1

R.T.: 5.117 min
Delta R.T.: 0.032 min
Response: 125528
Conc: 3.71 ng/ml




#4 AR-1016-2

R.T.: 0.000 min
Exp R.T.: 5.854 min
Response: 0
Conc: N.D.



#4 AR-1016-2


R.T.: 5.117 min
Delta R.T.: 0.012 min
Response: 125528
Conc: 2.49 ng/ml



#7 AR-1016-5


R.T.: 0.000 min
Exp R.T.: 6.310 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
FDA569G-1-2




#7 AR-1016-5

R.T.: 5.597 min
Delta R.T.: 0.055 min
Response: 143275
Conc: 4.74 ng/ml




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T.: 4.873 min
Response: 0
Conc: N.D.



#8 AR-1221-1


R.T.: 4.168 min
Delta R.T.: -0.028 min
Response: 96113
Conc: 7.12 ng/ml



#9 AR-1221-2


R.T.: 4.980 min
Delta R.T.: 0.021 min
Response: 122792
Conc: 9.54 ng/ml

Instrument :
ECD_D
ClientSampleId :
FDA569G-1-2




#9 AR-1221-2

R.T.: 4.228 min
Delta R.T.: -0.057 min
Response: 183332
Conc: 17.67 ng/ml




#10 AR-1221-3

R.T.: 4.980 min
Delta R.T.: -0.056 min
Response: 122792
Conc: 3.17 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T.: 4.362 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.980 min
Delta R.T.: -0.056 min
Response: 122792
Conc: 4.14 ng/ml

Instrument :
ECD_D
ClientSampleId :
FDA569G-1-2




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 4.362 min
Response: 0
Conc: N.D.




#12 AR-1232-2

R.T.: 0.000 min
Exp R.T. : 5.567 min
Response: 0
Conc: N.D.



#12 AR-1232-2


R.T.: 5.117 min
Delta R.T.: 0.013 min
Response: 125528
Conc: 5.25 ng/ml



#15 AR-1232-5


R.T.: 0.000 min
Exp R.T.: 6.107 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
FDA569G-1-2




#15 AR-1232-5

R.T.: 5.597 min
Delta R.T.: 0.055 min
Response: 143275
Conc: 11.16 ng/ml




#16 AR-1242-1

R.T.: 0.000 min
Exp R.T.: 5.835 min
Response: 0
Conc: N.D.



#16 AR-1242-1


R.T.: 5.117 min
Delta R.T.: 0.032 min
Response: 125528
Conc: 4.28 ng/ml



#17 AR-1242-2


R.T.: 0.000 min
Exp R.T.: 5.855 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
FDA569G-1-2




#17 AR-1242-2

R.T.: 5.117 min
Delta R.T.: 0.013 min
Response: 125528
Conc: 2.98 ng/ml




#21 AR-1248-1

R.T.: 0.000 min
Exp R.T.: 5.832 min
Response: 0
Conc: N.D.



#21 AR-1248-1


R.T.: 5.117 min
Delta R.T.: 0.031 min
Response: 125528
Conc: 5.46 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T.: 6.709 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
FDA569G-1-2




#24 AR-1248-4

R.T.: 5.597 min
Delta R.T.: 0.054 min
Response: 143275
Conc: 3.48 ng/ml




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T.: 7.552 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 6.624 min
Delta R.T.: -0.058 min
Response: 35237
Conc: 0.78 ng/ml



#30 AR-1254-5


R.T.: 0.000 min
Exp R.T.: 7.969 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
FDA569G-1-2




#30 AR-1254-5

R.T.: 7.093 min
Delta R.T.: -0.009 min
Response: 154628
Conc: 2.07 ng/ml




#31 AR-1260-1

R.T.: 0.000 min
Exp R.T.: 7.433 min
Response: 0
Conc: N.D.



#31 AR-1260-1


R.T.: 6.624 min
Delta R.T.: 0.040 min
Response: 35237
Conc: 0.62 ng/ml



#32 AR-1260-2


R.T.: 0.000 min
Exp R.T.: 7.687 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
FDA569G-1-2




#32 AR-1260-2

R.T.: 6.754 min
Delta R.T.: -0.017 min
Response: 154832
Conc: 2.35 ng/ml




#34 AR-1260-4

R.T.: 0.000 min
Exp R.T.: 8.280 min
Response: 0
Conc: N.D.



#34 AR-1260-4


R.T.: 7.442 min
Delta R.T.: 0.040 min
Response: 78347
Conc: 1.40 ng/ml



#35 AR-1260-5


R.T.: 0.000 min
Exp R.T.: 8.613 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
FDA569G-1-2




#35 AR-1260-5

R.T.: 7.597 min
Delta R.T.: -0.044 min
Response: 64111
Conc: 0.52 ng/ml




#36 AR-1262-1

R.T.: 0.000 min
Exp R.T.: 8.284 min
Response: 0
Conc: N.D.



#36 AR-1262-1


R.T.: 7.182 min
Delta R.T.: 0.040 min
Response: 213479
Conc: 2.67 ng/ml



#37 AR-1262-2


R.T.: 0.000 min
Exp R.T. : 8.615 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
FDA569G-1-2




#37 AR-1262-2

R.T.: 7.442 min
Delta R.T.: 0.038 min
Response: 78347
Conc: 1.12 ng/ml




#38 AR-1262-3

R.T.: 0.000 min
Exp R.T. : 8.945 min
Response: 0
Conc: N.D.



#38 AR-1262-3


R.T.: 7.893 min
Delta R.T.: -0.036 min
Response: 277459
Conc: 4.82 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T.: 9.724 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
FDA569G-1-2




#40 AR-1262-5

R.T.: 8.554 min
Delta R.T.: 0.043 min
Response: 88673
Conc: 1.82 ng/ml




#41 AR-1268-1

R.T.: 0.000 min
Exp R.T.: 8.941 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 7.893 min
Delta R.T.: -0.035 min
Response: 277459
Conc: 1.71 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 9.285 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
FDA569G-1-2




#43 AR-1268-3

R.T.: 8.231 min
Delta R.T.: 0.025 min
Response: 149415
Conc: 1.12 ng/ml



#44 AR-1268-4

R.T.: 0.000 min
Exp R.T.: 9.723 min
Response: 0
Conc: N.D.



#44 AR-1268-4

R.T.: 8.554 min
Delta R.T.: 0.043 min
Response: 88673
Conc: 1.66 ng/ml