

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD011625\
 Data File : PD087408.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Jan 2025 13:06
 Operator : AR\AJ
 Sample : INDA501
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA5055

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 16 13:16:12 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD011625CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Jan 16 13:16:00 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.555	2.882	154.9E6	884.5E6	76.027	71.516
2)	SA Decachlor...	9.086	8.085	481.9E6	2046.2E6	144.127	141.266
Target Compounds							
2)	A alpha-BHC	4.005	3.396	371.3E6	1492.6E6	89.377	75.375
3)	MA gamma-BHC...	4.336	3.733	350.0E6	1369.2E6	86.998	74.750
4)	MA Heptachlor	4.937	4.088	344.1E6	1352.5E6	84.191	72.466
9)	A Endosulfan I	6.082	5.254	281.3E6	1133.7E6	80.739	71.427
13)	MA Dieldrin	6.355	5.520	648.2E6	2426.5E6	170.026	145.381
14)	MA Endrin	6.582	5.796	549.4E6	2123.0E6	169.257	142.079
16)	A 4,4'-DDD	6.713	5.937	468.8E6	1933.2E6	170.495	144.597
17)	MA 4,4'-DDT	7.030	6.192	494.7E6	2061.1E6	170.674	149.564
20)	A Methoxychlor	7.502	6.764	1234.1E6	3928.6E6	778.937	637.829

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD011625\
Data File : PD087408.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Jan 2025 13:06
Operator : AR\AJ
Sample : INDA501
Misc :
ALS Vial : 17 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 16 13:16:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD011625CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jan 16 13:16:00 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Instrument :
ECD_D
ClientSampled :
INDA505