

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD011719\
 Data File : PD051091.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Jan 2019 1:01
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 1/18/2019 8:42:07 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 18 01:29:01 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD011019.M
 Quant Title : GC Extractables
 QLast Update : Thu Jan 10 03:41:46 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.226	4.045	81360491	1483.2E6	54.471	49.840
28)	SA Decachlor...	7.829	9.157	61108058	882.4E6	50.347	44.221
Target Compounds							
2)	A alpha-BHC	3.646	4.438	135.8E6	2279.2E6	56.403	50.015
3)	MA gamma-BHC...	3.914	4.725	127.2E6	2055.3E6	53.937	49.448
4)	MA Heptachlor	4.192	5.238	125.8E6	2008.6E6	52.525	49.232
5)	MB Aldrin	4.425	5.546	128.8E6	1915.8E6	54.283	48.942
6)	B beta-BHC	4.160	4.894	50476303	891.6E6	53.305	49.272
7)	B delta-BHC	4.349	5.113	126.4E6	1923.0E6	56.155	49.490
8)	B Heptachlor...	4.852	5.933	119.3E6	1659.7E6	53.098	47.600
9)	A Endosulfan I	5.174	6.291	111.9E6	1460.1E6	52.510	46.402
10)	B gamma-Chl...	5.069	6.167	123.9E6	1659.8E6	52.889	47.788
11)	B alpha-Chl...	5.125	6.241	121.0E6	1563.3E6	52.390	46.124
12)	B 4,4'-DDE	5.295	6.396	116.9E6	1479.2E6	54.142	44.867
13)	MA Dieldrin	5.409	6.550	123.9E6	1503.9E6	52.915	45.083
14)	MA Endrin	5.657	6.768	98576082	1138.6E6	45.009	40.428
15)	B Endosulfa...	5.926	6.976	103.8E6	1252.3E6	51.898	45.752m
16)	A 4,4'-DDD	5.797	6.888	96489291	1215.5E6	54.646	46.323
17)	MA 4,4'-DDT	6.028	7.188	95030347	1116.2E6	50.505	46.766
18)	B Endrin al...	6.093	7.102	86636625	1052.4E6	53.664	46.176
19)	B Endosulfa...	6.301	7.326	98909043	1237.9E6	50.494	45.856
20)	A Methoxychlor	6.575	7.646	48430025	558.1E6	50.675	45.847
21)	B Endrin ke...	6.776	7.799	113.3E6	1198.7E6	53.238	47.154

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD011719\
 Data File : PD051091.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Jan 2019 1:01
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 1/18/2019 8:42:07 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 18 01:29:01 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD011019.M
 Quant Title : GC Extractables
 QLast Update : Thu Jan 10 03:41:46 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

