

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD011719\
 Data File : PD051086.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 17 Jan 2019 23:51
 Operator : AJ\SJ
 Sample : K1167-08
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 D-10

Manual Integrations
 APPROVED

Sohil
 1/18/2019 8:42:03 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 18 00:06:32 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD011019.M
 Quant Title : GC Extractables
 QLast Update : Thu Jan 10 03:41:46 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.226	4.045	30592382	600.0E6	20.482	20.161
28)	SA Decachlor...	7.829	9.157	24221247	389.3E6	19.956	19.508m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD011719\
Data File : PD051086.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 17 Jan 2019 23:51
Operator : AJ\SJ
Sample : K1167-08
Misc :
ALS Vial : 50 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampled :
D-10


Manual Integrations
APPROVED

Sohil
1/18/2019 8:42:03 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 18 00:06:32 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD011019.M
Quant Title : GC Extractables
QLast Update : Thu Jan 10 03:41:46 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene

R.T.: 3.226 min
Delta R.T.: -0.001 min
Response: 30592382
Conc: 20.48 ng/ml

Instrument :
ECD_D
ClientSampleId :
D-10


Manual Integrations
APPROVED

Sohil
1/18/2019 8:42:03 AM




#1 Tetrachloro-m-xylene

R.T.: 4.045 min
Delta R.T.: -0.002 min
Response: 600003013
Conc: 20.16 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.829 min
Delta R.T.: -0.005 min
Response: 24221247
Conc: 19.96 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.157 min
Delta R.T.: -0.007 min
Response: 389263957
Conc: 19.51 ng/ml m