

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD011725\
 Data File : PD087426.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Jan 2025 17:42
 Operator : AR\AJ
 Sample : INDB301
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDB3068

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 17 00:55:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD011725CLP-TCLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Jan 17 00:54:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.555	2.882	43952937	265.8E6	20.000	20.000
27)	SA Decachlor...	9.087	8.084	135.1E6	594.2E6	40.000	40.000
Target Compounds							
5)	MB Aldrin	5.279	4.374	80926558	367.2E6	20.000	20.000
6)	B beta-BHC	4.520	4.028	36512943	173.5E6	20.000	20.000
7)	B delta-BHC	4.769	4.266	82639242	382.0E6	20.000	20.000
8)	B Heptachlo...	5.698	4.879	75120756	346.3E6	20.000	20.000
10)	B trans-Chl...	5.953	5.132	73820342	362.9E6	20.000	20.000
11)	B cis-Chlor...	6.035	5.197	74919228	351.1E6	20.000	20.000
12)	B 4,4'-DDE	6.204	5.383	141.1E6	667.2E6	40.000	40.000
15)	B Endosulfa...	6.794	6.088	129.0E6	602.9E6	40.000	40.000
18)	B Endrin al...	6.923	6.266	100.6E6	466.8E6	40.000	40.000
19)	B Endosulfa...	7.157	6.490	124.2E6	591.2E6	40.000	40.000
21)	B Endrin ke...	7.639	7.000	138.8E6	649.5E6	40.000	40.000


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD011725\
Data File : PD087426.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Jan 2025 17:42
Operator : AR\AJ
Sample : INDB301
Misc :
ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 17 00:55:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD011725CLP-TCLP.M
Quant Title : GC Extractables
QLast Update : Fri Jan 17 00:54:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.555 min
Delta R.T.: 0.000 min
Response: 43952937
Conc: 20.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDB3068




#1 Tetrachloro-m-xylene

R.T.: 2.882 min
Delta R.T.: 0.000 min
Response: 265812045
Conc: 20.00 ng/ml




#5 Aldrin

R.T.: 5.279 min
Delta R.T.: 0.000 min
Response: 80926558
Conc: 20.00 ng/ml



#5 Aldrin


R.T.: 4.374 min
Delta R.T.: 0.000 min
Response: 367228986
Conc: 20.00 ng/ml



#6 beta-BHC


R.T.: 4.520 min
Delta R.T.: 0.000 min
Response: 36512943
Conc: 20.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDB3068




#6 beta-BHC

R.T.: 4.028 min
Delta R.T.: 0.000 min
Response: 173472931
Conc: 20.00 ng/ml




#7 delta-BHC

R.T.: 4.769 min
Delta R.T.: 0.000 min
Response: 82639242
Conc: 20.00 ng/ml



#7 delta-BHC


R.T.: 4.266 min
Delta R.T.: 0.000 min
Response: 382040788
Conc: 20.00 ng/ml



#8 Heptachlor epoxide


R.T.: 5.698 min
Delta R.T.: 0.000 min
Response: 75120756
Conc: 20.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDB3068




#8 Heptachlor epoxide

R.T.: 4.879 min
Delta R.T.: 0.000 min
Response: 346286136
Conc: 20.00 ng/ml




#10 trans-Chlordane

R.T.: 5.953 min
Delta R.T.: 0.000 min
Response: 73820342
Conc: 20.00 ng/ml



#10 trans-Chlordane


R.T.: 5.132 min
Delta R.T.: 0.000 min
Response: 362933771
Conc: 20.00 ng/ml



#11 cis-Chlordane


R.T.: 6.035 min
Delta R.T.: 0.000 min
Response: 74919228
Conc: 20.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDB3068




#11 cis-Chlordane

R.T.: 5.197 min
Delta R.T.: 0.000 min
Response: 351137549
Conc: 20.00 ng/ml




#12 4,4'-DDE

R.T.: 6.204 min
Delta R.T.: 0.000 min
Response: 141102187
Conc: 40.00 ng/ml



#12 4,4'-DDE


R.T.: 5.383 min
Delta R.T.: 0.000 min
Response: 667245034
Conc: 40.00 ng/ml



#15 Endosulfan II


R.T.: 6.794 min
Delta R.T.: 0.000 min
Response: 128985513
Conc: 40.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDB3068




#15 Endosulfan II

R.T.: 6.088 min
Delta R.T.: 0.000 min
Response: 602890090
Conc: 40.00 ng/ml




#18 Endrin aldehyde

R.T.: 6.923 min
Delta R.T.: 0.000 min
Response: 100649862
Conc: 40.00 ng/ml



#18 Endrin aldehyde


R.T.: 6.266 min
Delta R.T.: 0.000 min
Response: 466830769
Conc: 40.00 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.157 min
Delta R.T.: 0.000 min
Response: 124239520
Conc: 40.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDB3068




#19 Endosulfan Sulfate

R.T.: 6.490 min
Delta R.T.: 0.000 min
Response: 591169386
Conc: 40.00 ng/ml




#21 Endrin ketone

R.T.: 7.639 min
Delta R.T.: 0.000 min
Response: 138750991
Conc: 40.00 ng/ml



#21 Endrin ketone


R.T.: 7.000 min
Delta R.T.: 0.000 min
Response: 649533469
Conc: 40.00 ng/ml



#27 Decachlorobiphenyl

R.T.: 9.087 min
Delta R.T.: 0.000 min
Response: 135111264
Conc: 40.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDB3068



#27 Decachlorobiphenyl

R.T.: 8.084 min
Delta R.T.: 0.000 min
Response: 594201204
Conc: 40.00 ng/ml