

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD011819\
 Data File : PD051099.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Jan 2019 16:00
 Operator : AJ\SJ
 Sample : K1148-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 T-1

Manual Integrations
 APPROVED

Sohil
 1/21/2019 9:59:47 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 18 22:05:26 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD011019.M
 Quant Title : GC Extractables
 QLast Update : Thu Jan 10 03:41:46 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.226	4.045	28578914	585.8E6	19.134	19.685
28)	SA Decachlor...	7.828	9.158	22424563	349.7E6	18.476	17.524
Target Compounds							
20)	A Methoxychlor	6.574	7.645	2548242	28929228	2.666	2.376m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD011819\
Data File : PD051099.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Jan 2019 16:00
Operator : AJ\SJ
Sample : K1148-01
Misc :
ALS Vial : 8 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
T-1


Manual Integrations
APPROVED

Sohil
1/21/2019 9:59:47 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 18 22:05:26 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD011019.M
Quant Title : GC Extractables
QLast Update : Thu Jan 10 03:41:46 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene

R.T.: 3.226 min
Delta R.T.: -0.001 min
Response: 28578914
Conc: 19.13 ng/ml

Instrument :
ECD_D
ClientSampleId :
T-1


Manual Integrations
APPROVED

Sohil
1/21/2019 9:59:47 AM




#1 Tetrachloro-m-xylene

R.T.: 4.045 min
Delta R.T.: -0.002 min
Response: 585829720
Conc: 19.68 ng/ml




#20 Methoxychlor

R.T.: 6.574 min
Delta R.T.: -0.004 min
Response: 2548242
Conc: 2.67 ng/ml



#20 Methoxychlor

R.T.: 7.645 min
Delta R.T.: -0.006 min
Response: 28929228
Conc: 2.38 ng/ml m




#28 Decachlorobiphenyl

R.T.: 7.828 min
Delta R.T.: -0.006 min
Response: 22424563
Conc: 18.48 ng/ml

Instrument :
ECD_D
ClientSampleId :
T-1

Manual Integrations
APPROVED

Sohil
1/21/2019 9:59:47 AM



#28 Decachlorobiphenyl

R.T.: 9.158 min
Delta R.T.: -0.006 min
Response: 349677950
Conc: 17.52 ng/ml