

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : DTX011820.M
 Title : GC Extractables
 Last Update : Mon Jan 20 06:13:56 2020
 Response Via : Initial Calibration

Calibration Files

500 =PD056983.D 1000 =PD056981.D 750 =PD056982.D
 250 =PD056984.D 100 =PD056985.D

Compound	500	1000	750	250	100	Avg	%RSD
1) SA Tetrachloro-m-xylene	8.188	8.097	8.349	8.125	8.312	8.214	E5 1.36
2) Toxaphene-1	0.992	0.958	1.008	1.017	1.060	1.007	E4 3.71
3) Toxaphene-2	1.516	1.478	1.550	1.571	1.566	1.536	E4 2.54
4) Toxaphene-3	1.522	1.473	1.617	1.575	1.356	1.509	E4 6.72
5) Toxaphene-4	6.758	6.754	7.299	6.796	5.653	6.652	E3 9.08
6) Toxaphene-5	6.901	6.146	7.227	5.438	7.502	6.643	E3 12.70
7) SA Decachlorobiphenyl	0.953	0.910	0.954	0.992	1.062	0.974	E6 5.87

Signal #2 Calibration Files

500 =PD056983.D 1000 =PD056981.D 750 =PD056982.D
 250 =PD056984.D 100 =PD056985.D

Compound	500	1000	750	250	100	Avg	%RSD
1) SA Tetrachloro-m-xylene	1.627	1.599	1.655	1.614	1.639	1.627	E6 1.33
2) Toxaphene-1	1.633	1.761	1.277	1.360	2.004	1.607	E4 18.46
3) Toxaphene-2	1.940	1.932	2.064	2.043	2.185	2.033	E4 5.10
4) Toxaphene-3	4.415	4.377	4.484	4.459	4.361	4.419	E4 1.18
5) Toxaphene-4	2.660	2.770	2.805	2.687	2.605	2.705	E4 3.01
6) Toxaphene-5	2.634	2.611	2.722	2.781	2.851	2.720	E4 3.69
7) SA Decachlorobiphenyl	1.626	1.532	1.617	1.704	1.809	1.658	E6 6.29

(#) = Out of Range