

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD011820\
 Data File : PD056988.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jan 2020 14:14
 Operator : AJ\MA
 Sample : PTOXICV500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICVPD011820TOX

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 20 06:17:26 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX011820.M
 Quant Title : GC Extractables
 QLast Update : Mon Jan 20 06:13:56 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	821.418	838.864 E3	-2.1	102	0.00
2 Toxaphene-1	10.071	10.004 E3	0.7	101	0.00
3 Toxaphene-2	15.362	14.999 E3	2.4	99	0.00
4 Toxaphene-3	15.088	16.072 E3	-6.5	106	0.00
5 Toxaphene-4	6.652	6.968 E3	-4.8	103	0.00
6 Toxaphene-5	6.643	6.674 E3	-0.5	97	0.00
7 SA Decachlorobiphenyl	974.192	986.745 E3	-1.3	104	0.00

Signal #2

9 SA Tetrachloro-m-xylene #2	1.627	1.658 E6	-1.9	102	0.00
10 Toxaphene-1 #2	16.070	13.823 E3	14.0	85	0.00
11 Toxaphene-2 #2	20.327	19.845 E3	2.4	102	0.00
12 Toxaphene-3 #2	44.192	44.999 E3	-1.8	102	0.00
13 Toxaphene-4 #2	27.052	27.952 E3	-3.3	105	0.00
14 Toxaphene-5 #2	27.200	27.533 E3	-1.2	105	0.00
15 SA Decachlorobiphenyl #2	1.658	1.689 E6	-1.9	104	0.00

Evaluate Continuing Calibration Report - Not Found

Signal #2

(#) = Out of Range

SPCC's out = 0 CCC's out = 0