

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD012023\
 Data File : PD073697.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jan 2023 17:04
 Operator : AR\AJ
 Sample : PB150361BSD
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PB150361BSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 20 21:25:55 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD011823.M
 Quant Title : GC Extractables
 QLast Update : Thu Jan 19 02:45:26 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.554	2.777	47218670	25111970	20.201	19.726
28)	SA Decachlor...	9.104	7.935	61959423	25105424	21.987	21.578
Target Compounds							
2)	A alpha-BHC	4.011	3.281	197.6E6	85690703	49.743	50.293
3)	MA gamma-BHC...	4.344	3.610	181.5E6	76262423	49.439	50.700
4)	MA Heptachlor	4.937	3.953	185.0E6	66372131	49.812	49.565
5)	MB Aldrin	5.281	4.235	186.3E6	81821950	51.683	50.656
6)	B beta-BHC	4.538	3.906	71628505	31080522	46.173	46.899
7)	B delta-BHC	4.787	4.137	182.0E6	77461647	54.091	50.677
8)	B Heptachlo...	5.708	4.738	164.1E6	67744336	51.421	50.347
9)	A Endosulfan I	6.095	5.108	149.6E6	64574550	48.715	48.027
10)	B gamma-Chl...	5.964	4.989	166.1E6	73077259	50.458	49.377
11)	B alpha-Chl...	6.044	5.053	163.4E6	72605363	49.524	48.814
12)	B 4,4'-DDE	6.216	5.244	151.7E6	67180419	50.221	50.347
13)	MA Dieldrin	6.371	5.374	164.3E6	68205990	51.531	52.033
14)	MA Endrin	6.602	5.649	134.6E6	57644216	49.654	51.837
15)	B Endosulfa...	6.819	5.943	136.8E6	60830280	51.662	52.231
16)	A 4,4'-DDD	6.733	5.798	123.3E6	52274359	49.983	50.733
17)	MA 4,4'-DDT	7.050	6.050	130.6E6	51671920	46.900	50.165
18)	B Endrin al...	6.950	6.123	112.7E6	48619650	51.939	55.124
19)	B Endosulfa...	7.185	6.346	130.7E6	56841850	48.808	50.497
20)	A Methoxychlor	7.524	6.627	71736452	26594988	46.308	50.812
21)	B Endrin ke...	7.672	6.853	148.6E6	65183334	49.150	50.874
22)	Mirex	8.149	7.039	111.3E6	53187491	48.436	50.644

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD012023\
Data File : PD073697.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jan 2023 17:04
Operator : AR\AJ
Sample : PB150361BSD
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB150361BSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 20 21:25:55 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD011823.M
Quant Title : GC Extractables
QLast Update : Thu Jan 19 02:45:26 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

