

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD012119\
 Data File : PD051137.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Jan 2019 15:17
 Operator : AJ\SJ
 Sample : PSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 21 16:18:03 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012119.M
 Quant Title : GC Extractables
 QLast Update : Mon Jan 21 16:16:54 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.225	4.045	90475492	1471.8E6	50.000	50.000
28)	SA Decachlor...	7.828	9.159	77754123	882.4E6	50.000	50.000
Target Compounds							
2)	A alpha-BHC	3.646	4.438	146.1E6	2220.5E6	50.000	50.000
3)	MA gamma-BHC...	3.914	4.725	139.1E6	2032.2E6	50.000	50.000
4)	MA Heptachlor	4.192	5.238	143.9E6	1947.5E6	50.000	50.000
5)	MB Aldrin	4.424	5.545	139.7E6	1833.0E6	50.000	50.000
6)	B beta-BHC	4.159	4.894	55685452	856.3E6	50.000	50.000
7)	B delta-BHC	4.348	5.114	122.0E6	1575.8E6	50.000	50.000
8)	B Heptachlo...	4.852	5.933	134.8E6	1661.8E6	50.000	50.000
9)	A Endosulfan I	5.174	6.292	130.5E6	1526.3E6	50.000	50.000
10)	B gamma-Chl...	5.069	6.168	145.5E6	1759.4E6	50.000	50.000
11)	B alpha-Chl...	5.125	6.241	141.2E6	1689.5E6	50.000	50.000
12)	B 4,4'-DDE	5.295	6.397	139.6E6	1662.3E6	50.000	50.000
13)	MA Dieldrin	5.408	6.550	147.9E6	1660.1E6	50.000	50.000
14)	MA Endrin	5.657	6.769	139.3E6	1360.6E6	50.000	50.000
15)	B Endosulfa...	5.926	6.978	131.7E6	1355.7E6	50.000	50.000
16)	A 4,4'-DDD	5.796	6.889	120.2E6	1319.3E6	50.000	50.000
17)	MA 4,4'-DDT	6.028	7.189	123.7E6	1174.7E6	50.000	50.000
18)	B Endrin al...	6.093	7.103	104.4E6	1068.2E6	50.000	50.000
19)	B Endosulfa...	6.301	7.326	117.6E6	1155.7E6	50.000	50.000
20)	A Methoxychlor	6.574	7.648	61434162	555.0E6	50.000	50.000
21)	B Endrin ke...	6.776	7.800	132.9E6	1160.9E6	50.000	50.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD012119\
Data File : PD051137.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Jan 2019 15:17
Operator : AJ\SJ
Sample : PSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 21 16:18:03 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012119.M
Quant Title : GC Extractables
QLast Update : Mon Jan 21 16:16:54 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

