

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD012125\
 Data File : PD087464.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jan 2025 10:04
 Operator : AR\AJ
 Sample : PEM047
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM036

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 22 03:17:21 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD011625CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Jan 16 13:39:07 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.558	2.881	43984022	262.4E6	21.582	21.218
2)	SA Decachlor...	9.085	8.080	73391666	316.6E6	21.952	21.857
Target Compounds							
2)	A alpha-BHC	4.008	3.394	41752910	212.2E6	10.050	10.717
3)	MA gamma-BHC...	4.339	3.731	41215785	194.5E6	10.246	10.616
6)	B beta-BHC	4.523	4.027	18664341	90592553	10.392	10.620
14)	MA Endrin	6.584	5.794	150.6E6	680.6E6	46.393	45.550
16)	A 4,4'-DDD	6.714	5.935	2702576	12833580	0.983	0.960
17)	MA 4,4'-DDT	7.031	6.189	310.2E6	1302.0E6	107.019	94.479
18)	B Endrin al...	6.923	6.263	3886193	24024415	1.542	2.084 #
20)	A Methoxychlor	7.502	6.761	405.1E6	1413.9E6	255.717	229.556
21)	B Endrin ke...	7.638	6.996	7800558	46650954	2.266	2.937 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD012125\
Data File : PD087464.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jan 2025 10:04
Operator : AR\AJ
Sample : PEM047
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 22 03:17:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD011625CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jan 16 13:39:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Instrument :
ECD_D
ClientSampled :
PEM036