

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD012320\
 Data File : PD057064.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jan 2020 17:59
 Operator : AJ\MA
 Sample : INDA401
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA401

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 24 08:19:12 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012320CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Jan 24 08:15:19 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.285	3.960	35535017	72668525	39.348	39.192
27)	SA Decachlor...	7.955	8.996	82431618	140.6E6	76.505	75.415
Target Compounds							
2)	A alpha-BHC	3.712	4.346	54109800	109.1E6	41.131	41.138
3)	MA gamma-BHC...	3.987	4.629	51641648	104.1E6	40.037	40.680
4)	MA Heptachlor	4.272	5.135	48870139	104.9E6	40.066	39.959
9)	A Endosulfan I	5.278	6.178	41442599	86207211	39.853	39.611
13)	MA Dieldrin	5.517	6.433	95959738	194.1E6	81.563	81.297
14)	MA Endrin	5.771	6.650	79359099	156.2E6	79.885	80.359
16)	A 4,4'-DDD	5.905	6.773	72365471	150.1E6	81.923	80.962
17)	MA 4,4'-DDT	6.140	7.071	73809601	155.1E6	85.008	81.754
20)	A Methoxychlor	6.688	7.527	183.6E6	394.3E6	388.573	386.376

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD012320\
Data File : PD057064.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jan 2020 17:59
Operator : AJ\MA
Sample : INDA401
Misc :
ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 24 08:19:12 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012320CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jan 24 08:15:19 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Instrument :
ECD_D
ClientSampled :
INDA401