

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD012323\
 Data File : PD073712.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jan 2023 14:42
 Operator : AR\AJ
 Sample : 01239-01MS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 BORING-3MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 23 20:57:21 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD011823.M
 Quant Title : GC Extractables
 QLast Update : Thu Jan 19 02:45:26 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.555	2.777	43137926	23272793	18.455	18.281
2)	SA Decachlor...	9.105	7.936	45046617	19233331	15.985	16.531
Target Compounds							
2)	A alpha-BHC	4.012	3.281	207.3E6	93144894	52.192	54.667
3)	MA gamma-BHC...	4.346	3.611	186.7E6	84088310	50.861	55.903
4)	MA Heptachlor	4.939	3.954	183.8E6	71535534	49.486	53.421
5)	MB Aldrin	5.283	4.235	186.8E6	89019827	51.814	55.112
6)	B beta-BHC	4.539	3.907	74415502	34900362	47.969	52.663
7)	B delta-BHC	4.789	4.138	169.5E6	84801214	50.374	55.479
8)	B Heptachlo...	5.710	4.739	163.3E6	75099275	51.158	55.813
9)	A Endosulfan I	6.096	5.110	148.6E6	68665635	48.393	51.070
10)	B gamma-Chl...	5.965	4.990	161.2E6	79214765	48.965m	53.524
11)	B alpha-Chl...	6.046	5.055	162.8E6	77637951	49.359	52.198
12)	B 4,4'-DDE	6.218	5.245	149.8E6	72084399	49.596	54.023
13)	MA Dieldrin	6.373	5.376	161.1E6	76472834	50.536	58.340
14)	MA Endrin	6.603	5.651	139.2E6	65405867	51.328	58.817
15)	B Endosulfa...	6.822	5.945	134.8E6	65658292	50.918	56.377
16)	A 4,4'-DDD	6.735	5.800	121.2E6	56343268	49.131	54.682
17)	MA 4,4'-DDT	7.051	6.052	127.0E6	52132873	45.598	50.612
18)	B Endrin al...	6.952	6.125	107.9E6	51629153	49.731	58.536
19)	B Endosulfa...	7.187	6.347	128.1E6	59245125	47.842	52.632
20)	A Methoxychlor	7.526	6.628	69800543	26903868	45.058	51.402
21)	B Endrin ke...	7.674	6.855	140.4E6	67170222	46.433	52.425
22)	Mirex	8.151	7.040	106.9E6	54106179	46.525	51.519

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD012323\
Data File : PD073712.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jan 2023 14:42
Operator : AR\AJ
Sample : 01239-01MS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
BORING-3MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 23 20:57:21 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD011823.M
Quant Title : GC Extractables
QLast Update : Thu Jan 19 02:45:26 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

