

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD012419CLP.M
 Title : GC Extractables
 Last Update : Fri Jan 25 03:14:31 2019
 Response Via : Initial Calibration

Calibration Files

3 =PD051208.D 1 =PD051204.D 2 =PD051206.D
 4 =PD051210.D 5 =PD051212.D

Compound			3	1	2	4	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene		1.362	1.413	1.384	1.353	1.330	1.368	E6	2.30
2) A	alpha-BHC							0.000		-1.00
3) MA	gamma-BHC (Lindane)							0.000		-1.00
4) MA	Heptachlor							0.000		-1.00
5) MB	Aldrin		2.033	1.962	1.985	2.087	2.127	2.039	E6	3.38
6) B	beta-BHC		0.938	1.018	0.974	0.898	0.858	0.937	E6	6.68
7) B	delta-BHC		2.123	2.005	2.064	2.212	2.255	2.132	E6	4.83
8) B	Heptachlor epoxide		2.006	2.009	2.001	2.021	2.027	2.013	E6	0.54
9) A	Endosulfan I							0.000		-1.00
10) B	trans-Chlordane		2.064	2.051	2.041	2.086	2.107	2.070	E6	1.29
11) B	cis-Chlordane		2.022	2.037	2.014	2.027	2.022	2.025	E6	0.42
12) B	4,4'-DDE		2.048	1.849	1.953	2.147	2.193	2.038	E6	6.90
13) MA	Dieldrin							0.000		-1.00
14) MA	Endrin							0.000		-1.00
15) B	Endosulfan II		1.970	1.909	1.939	1.999	1.999	1.963	E6	2.00
16) A	4,4'-DDD							0.000		-1.00
17) MA	4,4'-DDT							0.000		-1.00
18) B	Endrin aldehyde		1.551	1.552	1.562	1.552	1.529	1.549	E6	0.79
19) B	Endosulfan Sulfate		1.784	1.814	1.799	1.779	1.746	1.784	E6	1.41
20) A	Methoxychlor							0.000		-1.00
21) B	Endrin ketone		1.983	1.970	1.987	1.987	1.959	1.977	E6	0.62
22)	Toxaphene-1							0.000		-1.00
23)	Toxaphene-2							0.000		-1.00
24)	Toxaphene-3							0.000		-1.00
25)	Toxaphene-4							0.000		-1.00
26)	Toxaphene-5							0.000		-1.00
27) SA	Decachlorobiphenyl		1.181	1.380	1.280	1.111	1.030	1.197	E6	11.52

Signal #2 Calibration Files

3 =PD051208.D 1 =PD051204.D 2 =PD051206.D
 4 =PD051210.D 5 =PD051212.D

Compound			3	1	2	4	5	Avg		%RSD
1)	SA	Tetrachloro-m-xylene	2.835	3.107	3.038	2.633	2.418	2.806	E7	10.16
2)	A	alpha-BHC						0.000		-1.00
3)	MA	gamma-BHC (Lindane)						0.000		-1.00
4)	MA	Heptachlor						0.000		-1.00
5)	MB	Aldrin	3.724	4.072	3.957	3.466	3.188	3.682	E7	9.80
6)	B	beta-BHC	1.856	2.125	2.026	1.695	1.527	1.846	E7	13.14
7)	B	delta-BHC	3.757	3.506	3.749	3.570	3.398	3.596	E7	4.34
8)	B	Heptachlor epoxide	3.399	3.809	3.626	3.132	2.876	3.368	E7	11.10
9)	A	Endosulfan I						0.000		-1.00
10)	B	trans-Chlordane	3.433	3.734	3.631	3.177	2.927	3.380	E7	9.78
11)	B	cis-Chlordane	3.303	3.627	3.497	3.050	2.801	3.256	E7	10.26
12)	B	4,4'-DDE	3.132	3.397	3.297	2.941	2.717	3.097	E7	8.84
13)	MA	Dieldrin						0.000		-1.00
14)	MA	Endrin						0.000		-1.00
15)	B	Endosulfan II	2.677	2.995	2.868	2.481	2.263	2.657	E7	11.07
16)	A	4,4'-DDD						0.000		-1.00
17)	MA	4,4'-DDT						0.000		-1.00
18)	B	Endrin aldehyde	2.197	2.492	2.367	2.022	1.826	2.181	E7	12.19
19)	B	Endosulfan Sulfate	2.499	2.844	2.705	2.304	2.081	2.487	E7	12.28
20)	A	Methoxychlor						0.000		-1.00
21)	B	Endrin ketone	2.404	2.640	2.557	2.253	2.073	2.385	E7	9.60
22)		Toxaphene-1						0.000		-1.00
23)		Toxaphene-2						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
Method File : PD012419CLP.M
Title : GC Extractables
Last Update : Fri Jan 25 03:14:31 2019
Response Via : Initial Calibration

Calibration Files

3 =PD051208.D 1 =PD051204.D 2 =PD051206.D
4 =PD051210.D 5 =PD051212.D

	Compound	3	1	2	4	5	Avg	%RSD
24)	Toxaphene-3						0.000	-1.00
25)	Toxaphene-4						0.000	-1.00
26)	Toxaphene-5						0.000	-1.00
27) SA	Decachlorobiphenyl	1.877	2.197	2.065	1.665	1.528	1.866 E7	14.78

(#) = Out of Range