

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD012424\
 Data File : PD080864.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jan 2024 23:18
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 25 01:01:51 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012424.M
 Quant Title : GC Extractables
 QLast Update : Wed Jan 24 20:16:38 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.566	2.796	70956883	124.9E6	49.985	52.192
28)	SA Decachlor...	9.110	7.933	100.5E6	138.9E6	49.928	52.131
Target Compounds							
2)	A alpha-BHC	4.022	3.297	123.0E6	203.8E6	49.502	53.122
3)	MA gamma-BHC...	4.356	3.627	119.4E6	198.2E6	48.790	53.016
4)	MA Heptachlor	4.948	3.968	119.6E6	182.4E6	48.496	52.896
5)	MB Aldrin	5.292	4.248	120.5E6	196.0E6	49.092	53.828
6)	B beta-BHC	4.550	3.922	48929958	80557178	48.357	51.554
7)	B delta-BHC	4.799	4.153	123.4E6	202.8E6	49.195	54.029
8)	B Heptachlo...	5.718	4.749	110.2E6	174.3E6	49.231	52.026
9)	A Endosulfan I	6.105	5.119	102.7E6	159.6E6	49.503	52.116
10)	B gamma-Chl...	5.974	4.999	108.3E6	171.3E6	49.452	52.382
11)	B alpha-Chl...	6.054	5.064	107.9E6	170.6E6	49.550	52.166
12)	B 4,4'-DDE	6.225	5.252	101.9E6	163.0E6	50.083	53.201
13)	MA Dieldrin	6.380	5.384	110.9E6	179.6E6	50.167	52.814
14)	MA Endrin	6.611	5.659	93624509	155.7E6	50.127	52.540
15)	B Endosulfa...	6.828	5.951	95833607	151.5E6	49.490	52.669
16)	A 4,4'-DDD	6.742	5.805	83904840	139.1E6	51.064	52.984
17)	MA 4,4'-DDT	7.057	6.057	90770682	141.5E6	50.969	53.336
18)	B Endrin al...	6.958	6.131	75078427	117.1E6	49.356	51.950
19)	B Endosulfa...	7.193	6.353	94757033	145.8E6	49.905	52.277
20)	A Methoxychlor	7.530	6.631	50876603	81019125	49.429	51.761
21)	B Endrin ke...	7.679	6.859	106.1E6	171.3E6	50.738	53.090
22)	Mirex	8.156	7.042	82620155	130.0E6	50.092	51.590

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD012424\
Data File : PD080864.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jan 2024 23:18
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 25 01:01:51 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012424.M
Quant Title : GC Extractables
QLast Update : Wed Jan 24 20:16:38 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

