

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD012424\
 Data File : PD080864.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jan 2024 23:18
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 25 01:01:51 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012424.M
 Quant Title : GC Extractables
 QLast Update : Wed Jan 24 20:16:38 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.566	2.796	70956883	124.9E6	49.985	52.192
28)	SA Decachlor...	9.110	7.933	100.5E6	138.9E6	49.928	52.131
Target Compounds							
2)	A alpha-BHC	4.022	3.297	123.0E6	203.8E6	49.502	53.122
3)	MA gamma-BHC...	4.356	3.627	119.4E6	198.2E6	48.790	53.016
4)	MA Heptachlor	4.948	3.968	119.6E6	182.4E6	48.496	52.896
5)	MB Aldrin	5.292	4.248	120.5E6	196.0E6	49.092	53.828
6)	B beta-BHC	4.550	3.922	48929958	80557178	48.357	51.554
7)	B delta-BHC	4.799	4.153	123.4E6	202.8E6	49.195	54.029
8)	B Heptachlo...	5.718	4.749	110.2E6	174.3E6	49.231	52.026
9)	A Endosulfan I	6.105	5.119	102.7E6	159.6E6	49.503	52.116
10)	B gamma-Chl...	5.974	4.999	108.3E6	171.3E6	49.452	52.382
11)	B alpha-Chl...	6.054	5.064	107.9E6	170.6E6	49.550	52.166
12)	B 4,4'-DDE	6.225	5.252	101.9E6	163.0E6	50.083	53.201
13)	MA Dieldrin	6.380	5.384	110.9E6	179.6E6	50.167	52.814
14)	MA Endrin	6.611	5.659	93624509	155.7E6	50.127	52.540
15)	B Endosulfa...	6.828	5.951	95833607	151.5E6	49.490	52.669
16)	A 4,4'-DDD	6.742	5.805	83904840	139.1E6	51.064	52.984
17)	MA 4,4'-DDT	7.057	6.057	90770682	141.5E6	50.969	53.336
18)	B Endrin al...	6.958	6.131	75078427	117.1E6	49.356	51.950
19)	B Endosulfa...	7.193	6.353	94757033	145.8E6	49.905	52.277
20)	A Methoxychlor	7.530	6.631	50876603	81019125	49.429	51.761
21)	B Endrin ke...	7.679	6.859	106.1E6	171.3E6	50.738	53.090
22)	Mirex	8.156	7.042	82620155	130.0E6	50.092	51.590



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD012424\
Data File : PD080864.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jan 2024 23:18
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 25 01:01:51 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012424.M
Quant Title : GC Extractables
QLast Update : Wed Jan 24 20:16:38 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.566 min
Delta R.T.: 0.000 min
Response: 70956883
Conc: 49.98 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#1 Tetrachloro-m-xylene

R.T.: 2.796 min
Delta R.T.: 0.000 min
Response: 124850183
Conc: 52.19 ng/ml




#2 alpha-BHC

R.T.: 4.022 min
Delta R.T.: 0.000 min
Response: 122955309
Conc: 49.50 ng/ml



#2 alpha-BHC


R.T.: 3.297 min
Delta R.T.: 0.000 min
Response: 203835286
Conc: 53.12 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.356 min
Delta R.T.: 0.000 min
Response: 119419392
Conc: 48.79 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#3 gamma-BHC (Lindane)

R.T.: 3.627 min
Delta R.T.: 0.000 min
Response: 198221817
Conc: 53.02 ng/ml




#4 Heptachlor

R.T.: 4.948 min
Delta R.T.: 0.000 min
Response: 119608790
Conc: 48.50 ng/ml



#4 Heptachlor


R.T.: 3.968 min
Delta R.T.: 0.000 min
Response: 182390638
Conc: 52.90 ng/ml



#5 Aldrin


R.T.: 5.292 min
Delta R.T.: 0.000 min
Response: 120502448
Conc: 49.09 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#5 Aldrin

R.T.: 4.248 min
Delta R.T.: 0.000 min
Response: 196041472
Conc: 53.83 ng/ml




#6 beta-BHC

R.T.: 4.550 min
Delta R.T.: 0.000 min
Response: 48929958
Conc: 48.36 ng/ml



#6 beta-BHC


R.T.: 3.922 min
Delta R.T.: 0.000 min
Response: 80557178
Conc: 51.55 ng/ml



#7 delta-BHC


R.T.: 4.799 min
Delta R.T.: 0.000 min
Response: 123420210
Conc: 49.20 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#7 delta-BHC

R.T.: 4.153 min
Delta R.T.: 0.000 min
Response: 202771974
Conc: 54.03 ng/ml




#8 Heptachlor epoxide

R.T.: 5.718 min
Delta R.T.: 0.000 min
Response: 110238982
Conc: 49.23 ng/ml



#8 Heptachlor epoxide


R.T.: 4.749 min
Delta R.T.: 0.000 min
Response: 174260837
Conc: 52.03 ng/ml



#9 Endosulfan I


R.T.: 6.105 min
Delta R.T.: 0.000 min
Response: 102671519
Conc: 49.50 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#9 Endosulfan I

R.T.: 5.119 min
Delta R.T.: 0.000 min
Response: 159619131
Conc: 52.12 ng/ml




#10 gamma-Chlordane

R.T.: 5.974 min
Delta R.T.: 0.000 min
Response: 108255799
Conc: 49.45 ng/ml



#10 gamma-Chlordane


R.T.: 4.999 min
Delta R.T.: 0.000 min
Response: 171340966
Conc: 52.38 ng/ml



#11 alpha-Chlordane


R.T.: 6.054 min
Delta R.T.: 0.001 min
Response: 107890112
Conc: 49.55 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#11 alpha-Chlordane

R.T.: 5.064 min
Delta R.T.: 0.000 min
Response: 170557156
Conc: 52.17 ng/ml




#12 4,4'-DDE

R.T.: 6.225 min
Delta R.T.: 0.000 min
Response: 101913569
Conc: 50.08 ng/ml



#12 4,4'-DDE


R.T.: 5.252 min
Delta R.T.: 0.000 min
Response: 162969011
Conc: 53.20 ng/ml



#13 Dieldrin


R.T.: 6.380 min
Delta R.T.: 0.000 min
Response: 110852979
Conc: 50.17 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#13 Dieldrin

R.T.: 5.384 min
Delta R.T.: 0.000 min
Response: 179611293
Conc: 52.81 ng/ml




#14 Endrin

R.T.: 6.611 min
Delta R.T.: 0.000 min
Response: 93624509
Conc: 50.13 ng/ml



#14 Endrin


R.T.: 5.659 min
Delta R.T.: 0.000 min
Response: 155695858
Conc: 52.54 ng/ml



#15 Endosulfan II


R.T.: 6.828 min
Delta R.T.: 0.000 min
Response: 95833607
Conc: 49.49 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#15 Endosulfan II

R.T.: 5.951 min
Delta R.T.: 0.000 min
Response: 151470129
Conc: 52.67 ng/ml




#16 4,4'-DDD

R.T.: 6.742 min
Delta R.T.: 0.000 min
Response: 83904840
Conc: 51.06 ng/ml



#16 4,4'-DDD


R.T.: 5.805 min
Delta R.T.: 0.000 min
Response: 139096943
Conc: 52.98 ng/ml



#17 4,4'-DDT


R.T.: 7.057 min
Delta R.T.: 0.000 min
Response: 90770682
Conc: 50.97 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#17 4,4'-DDT

R.T.: 6.057 min
Delta R.T.: 0.000 min
Response: 141503011
Conc: 53.34 ng/ml




#18 Endrin aldehyde

R.T.: 6.958 min
Delta R.T.: 0.000 min
Response: 75078427
Conc: 49.36 ng/ml



#18 Endrin aldehyde


R.T.: 6.131 min
Delta R.T.: 0.000 min
Response: 117128635
Conc: 51.95 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.193 min
Delta R.T.: 0.000 min
Response: 94757033
Conc: 49.91 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#19 Endosulfan Sulfate

R.T.: 6.353 min
Delta R.T.: 0.000 min
Response: 145821305
Conc: 52.28 ng/ml




#20 Methoxychlor

R.T.: 7.530 min
Delta R.T.: 0.000 min
Response: 50876603
Conc: 49.43 ng/ml



#20 Methoxychlor


R.T.: 6.631 min
Delta R.T.: 0.000 min
Response: 81019125
Conc: 51.76 ng/ml



#21 Endrin ketone


R.T.: 7.679 min
Delta R.T.: 0.000 min
Response: 106105031
Conc: 50.74 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#21 Endrin ketone

R.T.: 6.859 min
Delta R.T.: 0.000 min
Response: 171322154
Conc: 53.09 ng/ml




#22 Mirex

R.T.: 8.156 min
Delta R.T.: 0.002 min
Response: 82620155
Conc: 50.09 ng/ml



#22 Mirex


R.T.: 7.042 min
Delta R.T.: 0.000 min
Response: 130004070
Conc: 51.59 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.110 min
Delta R.T.: 0.000 min
Response: 100477814
Conc: 49.93 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050



#28 Decachlorobiphenyl

R.T.: 7.933 min
Delta R.T.: 0.000 min
Response: 138856102
Conc: 52.13 ng/ml