

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD012521\
 Data File : PD061034.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jan 2021 15:52
 Operator : DD\AJ
 Sample : INDA401
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA4002

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 26 00:25:35 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012521CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 26 00:10:31 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.497	4.021	38926885	278.4E6	37.530	37.418
27)	SA Decachlor...	8.279	9.104	97008364	526.4E6	75.605	77.819
Target Compounds							
2)	A alpha-BHC	3.938	4.412	58952546	441.1E6	38.967	38.621
3)	MA gamma-BHC...	4.223	4.698	56287632	401.0E6	38.206	38.093
4)	MA Heptachlor	4.528	5.209	59288464	412.9E6	38.526	37.858
9)	A Endosulfan I	5.566	6.260	51179859	332.0E6	38.260	37.136
13)	MA Dieldrin	5.810	6.518	116.8E6	730.1E6	77.903	76.414
14)	MA Endrin	6.069	6.736	97883493	584.8E6	78.422	78.079
16)	A 4,4'-DDD	6.193	6.855	92406085	565.2E6	77.786	77.143
17)	MA 4,4'-DDT	6.435	7.155	99338177	576.3E6	78.978	78.176
20)	A Methoxychlor	6.981	7.612	242.6E6	1270.3E6	378.633	377.818

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD012521\
Data File : PD061034.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jan 2021 15:52
Operator : DD\AJ
Sample : INDA401
Misc :
ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 26 00:25:35 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012521CLP.M
Quant Title : GC Extractables
QLast Update : Tue Jan 26 00:10:31 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Instrument :
ECD_D
ClientSampled :
INDA4002