

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD012524\
 Data File : PD080881.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jan 2024 09:53
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 25 21:36:18 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012424.M
 Quant Title : GC Extractables
 QLast Update : Thu Jan 25 11:14:56 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.567	2.796	30120131	53112623	21.218	22.203
28)	SA Decachlor...	9.112	7.934	44533718	61921129	22.129	23.247
Target Compounds							
2)	A alpha-BHC	4.024	3.298	23264493	41984690	9.366	10.942
3)	MA gamma-BHC...	4.357	3.627	23693431	39331272	9.680	10.519
6)	B beta-BHC	4.551	3.922	11348064	18943379	11.215	12.123
14)	MA Endrin	6.612	5.659	89958880	150.6E6	48.164	50.811
16)	A 4,4'-DDD	6.744	5.807	1256874	2296835	0.765	0.875
17)	MA 4,4'-DDT	7.059	6.057	188.1E6	300.2E6	105.633	113.152
18)	B Endrin al...	6.960	6.133	2943290	5959891	1.935	2.643 #
20)	A Methoxychlor	7.532	6.632	248.4E6	382.3E6	241.350	244.238
21)	B Endrin ke...	7.680	6.860	6082926	10292501	2.909	3.189

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD012524\
Data File : PD080881.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jan 2024 09:53
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PEM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 25 21:36:18 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012424.M
Quant Title : GC Extractables
QLast Update : Thu Jan 25 11:14:56 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

