

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD012524\
 Data File : PD080903.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jan 2024 16:26
 Operator : AR\AJ
 Sample : P1240-04MSD
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 VNJ-232MSD

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 01/26/2024
 Supervised By :Ankita Jodhani 01/26/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 25 21:51:24 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012424.M
 Quant Title : GC Extractables
 QLast Update : Thu Jan 25 11:14:56 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.564	2.796	31931468	58370648	22.494m	24.401
28)	SA Decachlor...	9.110	7.933	41036376	55281567	20.391	20.754
Target Compounds							
2)	A alpha-BHC	4.022	3.298	122.7E6	195.4E6	49.412	50.933
3)	MA gamma-BHC...	4.354	3.627	116.9E6	186.6E6	47.758m	49.900
4)	MA Heptachlor	4.948	3.968	121.8E6	179.0E6	49.404	51.922
5)	MB Aldrin	5.292	4.248	116.4E6	180.1E6	47.424	49.449
6)	B beta-BHC	4.550	3.922	49434671	77855819	48.856	49.825
7)	B delta-BHC	4.799	4.153	115.3E6	179.2E6	45.957	47.756
8)	B Heptachlo...	5.718	4.749	108.8E6	164.2E6	48.606	49.016
9)	A Endosulfan I	6.105	5.119	100.2E6	154.3E6	48.311	50.387
10)	B gamma-Chl...	5.974	5.000	111.4E6	179.0E6	50.875	54.713
11)	B alpha-Chl...	6.054	5.064	111.6E6	177.0E6	51.241	54.128
12)	B 4,4'-DDE	6.225	5.252	98389796	156.7E6	48.352	51.155
13)	MA Dieldrin	6.380	5.384	107.3E6	172.9E6	48.562	50.828
14)	MA Endrin	6.611	5.659	90014287	153.8E6	48.194	51.891
15)	B Endosulfa...	6.829	5.952	93171534	147.5E6	48.115	51.292
16)	A 4,4'-DDD	6.742	5.805	80925131	130.0E6	49.251	49.512
17)	MA 4,4'-DDT	7.057	6.057	88476352	137.7E6	49.681	51.897
18)	B Endrin al...	6.959	6.130	71374225	111.0E6	46.921	49.225
19)	B Endosulfa...	7.193	6.353	89360704	139.6E6	47.063	50.031
20)	A Methoxychlor	7.532	6.631	47799966	75898051	46.440	48.489
21)	B Endrin ke...	7.680	6.859	101.1E6	161.4E6	48.322	50.006
22)	Mirex	8.156	7.043	77015950	123.6E6	46.694	49.033

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD012524\
Data File : PD080903.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jan 2024 16:26
Operator : AR\AJ
Sample : P1240-04MSD
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

VNJ-232MSD

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 01/26/2024
Supervised By :Ankita Jodhani 01/26/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 25 21:51:24 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012424.M
Quant Title : GC Extractables
QLast Update : Thu Jan 25 11:14:56 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

