

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD012723\
 Data File : PD073792.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Jan 2023 10:23
 Operator : AR\AJ
 Sample : INDA3105
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA3109

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 27 03:42:34 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012523CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Jan 25 02:23:06 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.555	2.776	49949330	27901576	21.092	21.981
27)	SA Decachlor...	9.106	7.935	120.4E6	50865427	43.643	42.263
Target Compounds							
2)	A alpha-BHC	4.011	3.280	81435815	38805127	21.381	21.898
3)	MA gamma-BHC...	4.345	3.611	78319027	36691269	21.503	21.825
4)	MA Heptachlor	4.938	3.953	78921882	31313937	21.691	20.763
9)	A Endosulfan I	6.096	5.109	65632401	30159666	21.387	21.016
13)	MA Dieldrin	6.373	5.375	143.8E6	65296525	42.804	41.952
14)	MA Endrin	6.603	5.651	115.0E6	53000813	41.091	41.180
16)	A 4,4'-DDD	6.735	5.799	109.3E6	51208428	42.974	42.456
17)	MA 4,4'-DDT	7.051	6.051	111.8E6	48142095	40.976	40.781
20)	A Methoxychlor	7.526	6.628	301.9E6	110.4E6	202.852	200.196

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD012723\
Data File : PD073792.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Jan 2023 10:23
Operator : AR\AJ
Sample : INDA3105
Misc :
ALS Vial : 7 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 27 03:42:34 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012523CLP.M
Quant Title : GC Extractables
QLast Update : Wed Jan 25 02:23:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

