

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD012920\
 Data File : PD057097.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Jan 2020 10:35
 Operator : AJ\MA
 Sample : INDA309
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA309

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 29 11:15:32 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012320CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Jan 24 08:33:42 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.285	3.960	17984010	38018889	19.890	20.545
27)	SA Decachlor...	7.958	8.996	42100164	75650715	38.671	40.482
Target Compounds							
2)	A alpha-BHC	3.712	4.347	26199895	54560310	19.806	20.658
3)	MA gamma-BHC...	3.987	4.629	26430803	52734685	19.903	20.589
4)	MA Heptachlor	4.272	5.136	23838681	53789528	19.392	20.461
9)	A Endosulfan I	5.278	6.178	21038849	44910264	19.958	20.628
13)	MA Dieldrin	5.518	6.434	47082656	97389406	39.745	40.844
14)	MA Endrin	5.771	6.651	36996175	74124115	36.921	38.135
16)	A 4,4'-DDD	5.906	6.773	35092242	74983489	39.582	40.599
17)	MA 4,4'-DDT	6.141	7.072	33207343	71894178	38.664	36.983
20)	A Methoxychlor	6.689	7.528	85106030	190.5E6	182.190	188.719

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD012920\
Data File : PD057097.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jan 2020 10:35
Operator : AJ\MA
Sample : INDA309
Misc :
ALS Vial : 8 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 29 11:15:32 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012320CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jan 24 08:33:42 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

