

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD012921\
 Data File : PD061097.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Jan 2021 15:51
 Operator : DD\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 30 00:18:22 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012821.M
 Quant Title : GC Extractables
 QLast Update : Fri Jan 29 01:48:44 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.496	4.020	44160257	344.9E6	50.008	51.036
28)	SA Decachlor...	8.277	9.105	43096550	306.2E6	50.043	50.218
Target Compounds							
2)	A alpha-BHC	3.936	4.411	71019966	566.0E6	50.284	51.020
3)	MA gamma-BHC...	4.221	4.697	67253607	521.4E6	50.246	51.284
4)	MA Heptachlor	4.527	5.208	63080093	512.8E6	48.404	50.478
5)	MB Aldrin	4.777	5.515	61501571	476.1E6	49.321	51.314
6)	B beta-BHC	4.471	4.865	28404094	207.1E6	49.110	50.779
7)	B delta-BHC	4.679	5.084	67373949	507.6E6	50.337	50.922
8)	B Heptachlo...	5.222	5.901	57704249	447.3E6	49.242	55.324
9)	A Endosulfan I	5.563	6.259	54419529	405.0E6	49.295	50.570
10)	B gamma-Chl...	5.450	6.135	57685249	440.2E6	49.061	51.364
11)	B alpha-Chl...	5.509	6.209	56982563	428.0E6	48.829	50.579
12)	B 4,4'-DDE	5.672	6.364	57647596	420.5E6	49.416	50.780
13)	MA Dieldrin	5.808	6.517	59906271	434.8E6	49.678	50.418
14)	MA Endrin	6.066	6.735	50680724	321.4E6	47.317	45.669
15)	B Endosulfa...	6.340	6.943	52210285	366.1E6	52.356	50.087
16)	A 4,4'-DDD	6.190	6.855	49955958	342.3E6	51.284	51.586
17)	MA 4,4'-DDT	6.432	7.154	47242302	327.4E6	48.487	49.326
18)	B Endrin al...	6.509	7.067	43561128	309.9E6	49.571	50.777
19)	B Endosulfa...	6.722	7.291	45682086	347.9E6	51.502	49.483
20)	A Methoxychlor	6.979	7.611	26686804	166.7E6	48.895	48.568
21)	B Endrin ke...	7.217	7.763	60452288	369.9E6	51.823	50.646
22)	Mirex	7.409	8.219	38432470	212.4E6	53.322	49.617

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD012921\
Data File : PD061097.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jan 2021 15:51
Operator : DD\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 30 00:18:22 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012821.M
Quant Title : GC Extractables
QLast Update : Fri Jan 29 01:48:44 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

