

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD020123\
 Data File : PD073807.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Feb 2023 13:28
 Operator : AR\AJ
 Sample : PIBLK016
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 20:42:11 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012523CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Jan 25 02:23:06 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.554	2.776	51883613	28716754	21.909	22.623
27)	SA Decachlor...	9.099	7.932	117.7E6	52584925	42.674	43.692
Target Compounds							
6)	B beta-BHC	4.566f	3.870f	362250	391682	0.216	0.479 #
7)	B delta-BHC	4.781	0.000	3480758	0	1.065	N.D. #
8)	B Heptachlo...	5.700	0.000	3994602	0	1.228	N.D. #
13)	MA Dieldrin	6.439f	0.000	2899488	0	0.863	N.D. #
14)	MA Endrin	6.627	5.665	775289	104771	0.277	<MDL #
15)	B Endosulfa...	6.793	0.000	474014	0	0.160	N.D. #
16)	A 4,4'-DDD	6.793f	0.000	474014	0	0.186	N.D. #
21)	B Endrin ke...	7.701	0.000	1487241	0	0.467	N.D. #
23)	Toxaphene-2	6.439	5.665	2899488	104771	98.514	12.649 #
24)	Toxaphene-3	6.627f	0.000	775289	0	36.871	N.D. #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD020123\
Data File : PD073807.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Feb 2023 13:28
Operator : AR\AJ
Sample : PIBLK016
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 20:42:11 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012523CLP.M
Quant Title : GC Extractables
QLast Update : Wed Jan 25 02:23:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.554 min
Delta R.T.: 0.000 min
Response: 51883613
Conc: 21.91 ng/ml

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 2.776 min
Delta R.T.: 0.000 min
Response: 28716754
Conc: 22.62 ng/ml




#6 beta-BHC

R.T.: 4.566 min
Delta R.T.: 0.029 min
Response: 362250
Conc: 0.22 ng/ml



#6 beta-BHC


R.T.: 3.870 min
Delta R.T.: -0.036 min
Response: 391682
Conc: 0.48 ng/ml



#7 delta-BHC


R.T.: 4.781 min
Delta R.T.: -0.004 min
Response: 3480758
Conc: 1.07 ng/ml

Instrument :
ECD_D
ClientSampleId :




#7 delta-BHC

R.T.: 0.000 min
Exp R.T. : 4.137 min
Response: 0
Conc: N.D.




#8 Heptachlor epoxide

R.T.: 5.700 min
Delta R.T.: -0.006 min
Response: 3994602
Conc: 1.23 ng/ml



#8 Heptachlor epoxide


R.T.: 0.000 min
Exp R.T. : 4.737 min
Response: 0
Conc: N.D.



#13 Dieldrin


R.T.: 6.439 min
Delta R.T.: 0.070 min
Response: 2899488
Conc: 0.86 ng/ml

Instrument :
ECD_D
ClientSampleId :




#13 Dieldrin

R.T.: 0.000 min
Exp R.T. : 5.373 min
Response: 0
Conc: N.D.




#14 Endrin

R.T.: 6.627 min
Delta R.T.: 0.027 min
Response: 775289
Conc: 0.28 ng/ml



#14 Endrin


R.T.: 5.665 min
Delta R.T.: 0.016 min
Response: 104771
Conc: N.D.



#15 Endosulfan II


R.T.: 6.793 min
Delta R.T.: -0.024 min
Response: 474014
Conc: 0.16 ng/ml

Instrument :
ECD_D
ClientSampleId :




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T. : 5.942 min
Response: 0
Conc: N.D.




#16 4,4'-DDD

R.T.: 6.793 min
Delta R.T.: 0.061 min
Response: 474014
Conc: 0.19 ng/ml



#16 4,4'-DDD


R.T.: 0.000 min
Exp R.T. : 5.797 min
Response: 0
Conc: N.D.



#21 Endrin ketone


R.T.: 7.701 min
Delta R.T.: 0.032 min
Response: 1487241
Conc: 0.47 ng/ml

Instrument :
ECD_D
ClientSampleId :




#21 Endrin ketone

R.T.: 0.000 min
Exp R.T. : 6.851 min
Response: 0
Conc: N.D.




#23 Toxaphene-2

R.T.: 6.439 min
Delta R.T.: -0.026 min
Response: 2899488
Conc: 98.51 ng/ml



#23 Toxaphene-2


R.T.: 5.665 min
Delta R.T.: -0.033 min
Response: 104771
Conc: 12.65 ng/ml



#24 Toxaphene-3


R.T.: 6.627 min
Delta R.T.: -0.038 min
Response: 775289
Conc: 36.87 ng/ml

Instrument :
ECD_D
ClientSampleId :




#24 Toxaphene-3

R.T.: 0.000 min
Exp R.T. : 5.980 min
Response: 0
Conc: N.D.



#27 Decachlorobiphenyl

R.T.: 9.099 min
Delta R.T.: 0.000 min
Response: 117735040
Conc: 42.67 ng/ml



#27 Decachlorobiphenyl

R.T.: 7.932 min
Delta R.T.: 0.000 min
Response: 52584925
Conc: 43.69 ng/ml