

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD020322\
 Data File : PD067920.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Feb 2022 17:39
 Operator : AR\AJ
 Sample : INDB368
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDB3127

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 04 05:59:32 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD020322CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Feb 04 02:10:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.435	3.958	38829291	152.4E6	19.198	19.411
27)	SA Decachlor...	8.157	9.014	32634995	188.3E6	35.715	38.455
Target Compounds							
5)	MB Aldrin	4.691	5.440	56208843	157.2E6	19.089	19.200
6)	B beta-BHC	4.403	4.816	24896852	86427049	18.699	19.195
7)	B delta-BHC	4.604	5.030	54687591	170.3E6	19.435	20.827
8)	B Heptachlo...	5.132	5.829	50934630	130.9E6	19.108	19.293
10)	B trans-Chl...	5.355	6.063	51952133	141.0E6	19.094	19.352
11)	B cis-Chlor...	5.414	6.134	51184140	140.5E6	19.007	19.205
12)	B 4,4'-DDE	5.578	6.295	95065376	261.1E6	38.875	38.513
15)	B Endosulfa...	6.239	6.878	80428912	237.7E6	37.795	38.493
18)	B Endrin al...	6.408	7.003	58775374	184.6E6	37.680	38.462
19)	B Endosulfa...	6.620	7.226	67647732	226.7E6	37.409	38.582
21)	B Endrin ke...	7.112	7.698	70465015	259.8E6	37.266	38.850

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD020322\
Data File : PD067920.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Feb 2022 17:39
Operator : AR\AJ
Sample : INDB368
Misc :
ALS Vial : 33 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 04 05:59:32 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD020322CLP.M
Quant Title : GC Extractables
QLast Update : Fri Feb 04 02:10:59 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

