

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD020323\
 Data File : PD073821.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Feb 2023 18:39
 Operator : AR\AJ
 Sample : INDA3108
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA3116

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 04 00:02:10 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012523CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Jan 25 02:23:06 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.554	2.775	49492656	26744950	20.899	21.070
27) SA Decachlor...	9.101	7.931	128.1E6	53234964	46.433	44.232
Target Compounds						
2) A alpha-BHC	4.010	3.279	77978454	37135210	20.473	20.955
3) MA gamma-BHC...	4.343	3.609	75284457	35147729	20.670	20.907
4) MA Heptachlor	4.936	3.952	76907969	30881652	21.137	20.476
9) A Endosulfan I	6.094	5.107	63605898	29261958	20.727	20.391
13) MA Dieldrin	6.370	5.373	138.1E6	62941670	41.100	40.439
14) MA Endrin	6.601	5.648	111.3E6	51094818	39.785	39.699
16) A 4,4'-DDD	6.733	5.797	105.8E6	49360075	41.581	40.923
17) MA 4,4'-DDT	7.048	6.048	107.9E6	46655418	39.543	39.522
20) A Methoxychlor	7.523	6.624	291.1E6	106.6E6	195.562	193.214

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD073823\
Data File : PD073821.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Feb 2023 18:39
Operator : AR\AJ
Sample : INDA3108
Misc :
ALS Vial : 11 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 04 00:02:10 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012523CLP.M
Quant Title : GC Extractables
QLast Update : Wed Jan 25 02:23:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

