

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020420\
 Data File : PS008705.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Feb 2020 15:05
 Operator : DD\AJ
 Sample : HSTDICV750
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 05 02:38:06 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020420.M
 Quant Title : 8080.M
 QLast Update : Wed Feb 05 02:36:36 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	Dalapon	2.536	2.570 E6	-1.3	99	0.00
2 T	3,5-DICHLORO BENZOIC ACID	2.824	2.829 E6	-0.2	101	0.00
3 T	4-Nitrophenol	1.085	1.086 E6	-0.1	100	0.00
4 S	2,4-DCAA	2.146	2.151 E6	-0.2	101	0.00
5 T	DICAMBA	8.066	8.132 E6	-0.8	100	0.00
6 T	MCP P	6.147	6.291 E6	-2.3	100	0.00
7 T	MCPA	8.590	8.759 E6	-2.0	100	0.00
8 T	DICHLORPROP	1.978	1.971 E6	0.4	100	0.00
9 T	2,4-D	2.140	2.154 E6	-0.7	100	0.00
10 T	Pentachlorophenol	28.429	28.335 E6	0.3	100	0.00
11 T	2,4,5-TP (SILVEX)	11.178	11.128 E6	0.4	100	0.00
12 T	2,4,5-T	11.006	10.981 E6	0.2	99	0.00
13 T	2,4-DB	1.627	1.625 E6	0.1	100	0.00
14 T	DINOSEB	8.439	8.374 E6	0.8	99	0.00
15 T	Picloram	15.144	15.457 E6	-2.1	100	0.00
16 T	DCPA	14.019	13.955 E6	0.5	99	0.00

Signal #2

1 T	Dalapon	1.199	1.223 E6	-2.0	99	0.00
2 T	3,5-DICHLORO BENZOIC ACID	1.376	1.382 E6	-0.4	101	0.00
3 T	4-Nitrophenol	568.578	571.126 E3	-0.4	101	0.00
4 S	2,4-DCAA	926.824	929.923 E3	-0.3	101	0.00
5 T	DICAMBA	3.891	3.951 E6	-1.5	100	0.00
6 T	MCP P	2.935	3.008 E6	-2.5	100	0.00
7 T	MCPA	4.174	4.235 E6	-1.5	101	0.00
8 T	DICHLORPROP	952.701	954.672 E3	-0.2	101	0.00
9 T	2,4-D	1.141	1.150 E6	-0.8	100	0.00
10 T	Pentachlorophenol	13.189	13.236 E6	-0.4	101	0.00
11 T	2,4,5-TP (SILVEX)	5.125	5.157 E6	-0.6	101	0.00
12 T	2,4,5-T	5.168	5.185 E6	-0.3	101	0.00
13 T	2,4-DB	699.268	700.764 E3	-0.2	101	0.00
14 T	DINOSEB	3.673	3.712 E6	-1.1	100	0.00
15 T	Picloram	7.548	7.768 E6	-2.9	101	0.00
16 T	DCPA	6.504	6.567 E6	-1.0	101	0.00

Evaluate Continuing Calibration Report - Not Found

Signal #2

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020420\
Data File : PS008705.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Feb 2020 15:05
Operator : DD\AJ
Sample : HSTDICV750
Misc :
ALS Vial : 8 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 05 02:38:06 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020420.M
Quant Title : 8080.M
QLast Update : Wed Feb 05 02:36:36 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0