

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD020422\
 Data File : PD067950.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Feb 2022 17:02
 Operator : AR\AJ
 Sample : INDA369
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA3128

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 02/07/2022
 Supervised By : Ankita Jodhani 02/08/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 07 00:37:58 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD020322CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Feb 04 02:10:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.434	3.959	39468673	158.5E6	19.515	20.199
27)	SA Decachlor...	8.155	9.015	30020706	175.6E6	32.854	35.847
Target Compounds							
2)	A alpha-BHC	3.871	4.353	57085988	192.0E6	19.138	19.162
3)	MA gamma-BHC...	4.152	4.638	53111467	202.2E6	18.451	21.259
4)	MA Heptachlor	4.443	5.138	51805958	152.2E6	17.651m	17.862
9)	A Endosulfan I	5.466	6.187	44564288	129.4E6	18.157	18.576
13)	MA Dieldrin	5.707	6.446	99066309	277.1E6	37.268m	38.607
14)	MA Endrin	5.966	6.662	75404490	208.1E6	35.026	34.960m
16)	A 4,4'-DDD	6.091	6.791	74112437	220.3E6	37.620	38.052
17)	MA 4,4'-DDT	6.329	7.088	62878688	191.9E6	32.833	33.334
20)	A Methoxychlor	6.876	7.551	141.2E6	457.3E6	160.759	163.408

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD020422\
Data File : PD067950.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Feb 2022 17:02
Operator : AR\AJ
Sample : INDA369
Misc :
ALS Vial : 17 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 07 00:37:58 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD020322CLP.M
Quant Title : GC Extractables
QLast Update : Fri Feb 04 02:10:59 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Manual Integrations
APPROVED
Reviewed By: Abdul Mirza 02/07/2022
Supervised By: Ankita Jodhani 02/08/2022

Instrument :
ECD_D
ClientSampled :
INDA3128