

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD020422\
 Data File : PD067941.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Feb 2022 14:09
 Operator : AR\AJ
 Sample : N1307-11
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 C0Q34

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 02/07/2022
 Supervised By : Ankita Jodhani 02/08/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 07 00:35:02 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD020322CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Feb 04 02:10:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.431	3.957	44468988	279.2E6	21.987	35.567m#
27)	SA Decachlor...	8.152	9.013	30730226	184.0E6	33.630	37.561
Target Compounds							
8)	B Heptachlo...	5.125	5.827	14468444	62626941	5.428m	9.229m#
9)	A Endosulfan I	5.443	6.171	9643727	24477905	3.929	3.514
10)	B trans-Chl...	5.348	6.063	2412195	16182006	0.887m	2.220m#
11)	B cis-Chlor...	5.408	6.136	5988329	22270712	2.224	3.045 #
12)	B 4,4'-DDE	5.579	6.288	2683723	2711642	1.097m	0.400 #
13)	MA Dieldrin	5.704	6.444	93072936	261.8E6	35.014	36.472
17)	MA 4,4'-DDT	6.330	7.077	11922555	5374496	6.226	0.934m#
21)	B Endrin ke...	7.106	7.696	3771186	18571406	1.994	2.777m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD020422\
Data File : PD067941.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Feb 2022 14:09
Operator : AR\AJ
Sample : N1307-11
Misc :
ALS Vial : 9 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 07 00:35:02 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD020322CLP.M
Quant Title : GC Extractables
QLast Update : Fri Feb 04 02:10:59 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

