

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD020425\  
Data File : PD087700.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 04 Feb 2025 13:15  
Operator : AR\AJ  
Sample : Q1226-22  
Misc :  
ALS Vial : 7 Sample Multiplier: 1



Instrument :  
ECD\_D  
ClientSampleId :  
DCZL9

## Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 02/05/2025  
Supervised By :Ankita Jodhani 02/05/2025

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Feb 05 08:01:37 2025  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD011625CLP.M  
Quant Title : GC Extractables  
QLast Update : Thu Jan 16 13:39:07 2025  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

3.559min 12.974 ng/ml

response 26440988

(1) Tetrachloro-m-xylene #2 (SA)

2.883min 19.475 ng/ml

response 240868032

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD020425\  
Data File : PD087700.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 04 Feb 2025 13:15  
Operator : AR\AJ  
Sample : Q1226-22  
Misc :  
ALS Vial : 7 Sample Multiplier: 1


Instrument :  
ECD\_D  
ClientSampleId :  
DCZL9

## Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 02/05/2025  
Supervised By :Ankita Jodhani 02/05/2025

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Feb 05 08:01:37 2025  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD011625CLP.M  
Quant Title : GC Extractables  
QLast Update : Thu Jan 16 13:39:07 2025  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

3.558min 21.071 ng/ml m

response 42941626

(1) Tetrachloro-m-xylene #2 (SA)

2.883min 19.475 ng/ml

response 240868032

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD020425\  
Data File : PD087700.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 04 Feb 2025 13:15  
Operator : AR\AJ  
Sample : Q1226-22  
Misc :  
ALS Vial : 7 Sample Multiplier: 1


Instrument :  
ECD\_D  
ClientSampleId :  
DCZL9

## Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 02/05/2025  
Supervised By :Ankita Jodhani 02/05/2025

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Feb 05 08:01:37 2025  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD011625CLP.M  
Quant Title : GC Extractables  
QLast Update : Thu Jan 16 13:39:07 2025  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(3) gamma-BHC (Lindane) (MA)

4.341min -1128.169 ng/ml

response -4538176608

(3) gamma-BHC (Lindane) #2 (MA)

3.733min 82.574 ng/ml

response 1512487047

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD020425\  
Data File : PD087700.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 04 Feb 2025 13:15  
Operator : AR\AJ  
Sample : Q1226-22  
Misc :  
ALS Vial : 7 Sample Multiplier: 1


Instrument :  
ECD\_D  
ClientSampleId :  
DCZL9

## Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 02/05/2025  
Supervised By :Ankita Jodhani 02/05/2025

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Feb 05 08:01:37 2025  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD011625CLP.M  
Quant Title : GC Extractables  
QLast Update : Thu Jan 16 13:39:07 2025  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(3) gamma-BHC (Lindane) (MA)

4.340min 111.806 ng/ml m

response 449751287

(3) gamma-BHC (Lindane) #2 (MA)

3.733min 82.574 ng/ml

response 1512487047

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD020425\  
Data File : PD087700.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 04 Feb 2025 13:15  
Operator : AR\AJ  
Sample : Q1226-22  
Misc :  
ALS Vial : 7 Sample Multiplier: 1



Instrument :  
ECD\_D  
ClientSampleId :  
DCZL9

## Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 02/05/2025  
Supervised By :Ankita Jodhani 02/05/2025

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Feb 05 08:01:37 2025  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD011625CLP.M  
Quant Title : GC Extractables  
QLast Update : Thu Jan 16 13:39:07 2025  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(11) cis-Chlordane (B)

6.079min 0.270 ng/ml

response 1002250

(11) cis-Chlordane #2 (B)

5.200min 1.059 ng/ml

response 18321127

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD020425\  
Data File : PD087700.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 04 Feb 2025 13:15  
Operator : AR\AJ  
Sample : Q1226-22  
Misc :  
ALS Vial : 7 Sample Multiplier: 1



Instrument :  
ECD\_D  
ClientSampleId :  
DCZL9

## Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 02/05/2025  
Supervised By :Ankita Jodhani 02/05/2025

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Feb 05 08:01:37 2025  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD011625CLP.M  
Quant Title : GC Extractables  
QLast Update : Thu Jan 16 13:39:07 2025  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(11) cis-Chlordane (B)

6.041min 0.695 ng/ml m

response 2582777

(11) cis-Chlordane #2 (B)

5.199min 0.598 ng/ml m

response 10343003

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD020425\  
Data File : PD087700.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 04 Feb 2025 13:15  
Operator : AR\AJ  
Sample : Q1226-22  
Misc :  
ALS Vial : 7 Sample Multiplier: 1


Instrument :  
ECD\_D  
ClientSampleId :  
DCZL9

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 02/05/2025  
Supervised By :Ankita Jodhani 02/05/2025

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Feb 05 08:01:37 2025  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD011625CLP.M  
Quant Title : GC Extractables  
QLast Update : Thu Jan 16 13:39:07 2025  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(20) Methoxychlor (A)

7.501min 1.051 ng/ml

response 1665702

(20) Methoxychlor #2 (A)

6.773min 0.935 ng/ml

response 5761525

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD020425\  
Data File : PD087700.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 04 Feb 2025 13:15  
Operator : AR\AJ  
Sample : Q1226-22  
Misc :  
ALS Vial : 7 Sample Multiplier: 1


Instrument :  
ECD\_D  
ClientSampleId :  
DCZL9

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 02/05/2025  
Supervised By :Ankita Jodhani 02/05/2025

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Feb 05 08:01:37 2025  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD011625CLP.M  
Quant Title : GC Extractables  
QLast Update : Thu Jan 16 13:39:07 2025  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(20) Methoxychlor (A)  
7.499min 0.950 ng/ml m  
response 1504421

(20) Methoxychlor #2 (A)  
6.773min 0.756 ng/ml m  
response 4655314