

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD020624\
 Data File : PD081116.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Feb 2024 13:07
 Operator : AR\AJ
 Sample : PB158813BS
 Misc : TOX BS
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PB158813BS

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 02/07/2024
 Supervised By :Ankita Jodhani 02/07/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 06 13:19:52 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX012424.M
 Quant Title : GC Extractables
 QLast Update : Wed Jan 24 20:34:35 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.573	2.793	33197232	64183461	22.629	24.985
7) SA Decachlor...	9.121	7.934	48938630	68100398	23.516	24.310
Target Compounds						
2) Toxaphene-1	6.281	5.023	4937421	5544208	253.232	265.210
3) Toxaphene-2	6.683	5.706	3667329	5175966	226.975	231.247
4) Toxaphene-3	7.101	6.746	14439429	22904427	226.570	232.556
5) Toxaphene-4	7.195	7.061	10859813	9730707	232.754	242.217m
6) Toxaphene-5	7.614	7.193	7994635	6104114	229.893m	204.488

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD020624\
Data File : PD081116.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Feb 2024 13:07
Operator : AR\AJ
Sample : PB158813BS
Misc : TOX BS
ALS Vial : 8 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 06 13:19:52 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX012424.M
Quant Title : GC Extractables
QLast Update : Wed Jan 24 20:34:35 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x0.25µm

