

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD020819\
 Data File : PD051249.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 08 Feb 2019 11:42
 Operator : AJ\SJ
 Sample : PTOXICC500
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PTOXICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 11 11:24:00 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX020819.M
 Quant Title : GC Extractables
 QLast Update : Fri Feb 08 23:29:43 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.225	4.045	67882178	1372.9E6	50.000	50.000
7)	SA Decachlor...	7.828	9.159	52611846	841.4E6	50.000	50.000
Target Compounds							
2)	Toxaphene-1	5.526	6.384	5302843	35589721	500.000	500.000
3)	Toxaphene-2	5.957	6.740	14480324	80654137	500.000	500.000
4)	Toxaphene-3	6.333	7.225	7179702	305.8E6	500.000	500.000
5)	Toxaphene-4	6.542	7.312	11136611	185.5E6	500.000	500.000
6)	Toxaphene-5	6.671	7.453	16921472	83024908	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD020819\
Data File : PD051249.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Feb 2019 11:42
Operator : AJ\SJ
Sample : PTOXICC500
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampled :
PTOXICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 11 11:24:00 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX020819.M
Quant Title : GC Extractables
QLast Update : Fri Feb 08 23:29:43 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

