

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD020823\
 Data File : PD073858.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Feb 2023 19:18
 Operator : AR\AJ
 Sample : 01463-12
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 COMP-6

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 08 03:13:47 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD011823.M
 Quant Title : GC Extractables
 QLast Update : Thu Jan 19 02:45:26 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.554	2.776	23537674	13126891	10.070	10.311
28)	SA Decachlor...	9.101	7.931	38695826	16295297	13.731	14.006
Target Compounds							
17)	MA 4,4'-DDT	7.048	6.045	1389522	1114944	0.499	1.082 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD020823\
Data File : PD073858.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Feb 2023 19:18
Operator : AR\AJ
Sample : 01463-12
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
COMP-6

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 08 03:13:47 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD011823.M
Quant Title : GC Extractables
QLast Update : Thu Jan 19 02:45:26 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.554 min
Delta R.T.: -0.001 min
Response: 23537674
Conc: 10.07 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-6




#1 Tetrachloro-m-xylene

R.T.: 2.776 min
Delta R.T.: -0.001 min
Response: 13126891
Conc: 10.31 ng/ml




#17 4,4'-DDT

R.T.: 7.048 min
Delta R.T.: -0.002 min
Response: 1389522
Conc: 0.50 ng/ml



#17 4,4'-DDT


R.T.: 6.045 min
Delta R.T.: -0.005 min
Response: 1114944
Conc: 1.08 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.101 min
Delta R.T.: -0.001 min
Response: 38695826
Conc: 13.73 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-6



#28 Decachlorobiphenyl

R.T.: 7.931 min
Delta R.T.: -0.003 min
Response: 16295297
Conc: 14.01 ng/ml