

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD020921\
 Data File : PD061143.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Feb 2021 11:41
 Operator : DD\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 09 11:52:58 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012821.M
 Quant Title : GC Extractables
 QLast Update : Fri Jan 29 01:48:44 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.495	4.020	22561319	141.5E6	25.549	20.937
28)	SA Decachlor...	8.273	9.103	25783879	147.3E6	29.940	24.150
Target Compounds							
2)	A alpha-BHC	3.935	4.410	18052752	108.9E6	12.782	9.818
3)	MA gamma-BHC...	4.220	4.696	16839660	102.9E6	12.581	10.125
6)	B beta-BHC	4.470	4.865	8094927	44454012	13.996	10.900
12)	B 4,4'-DDE	0.000	6.363	0	2173612	N.D.	0.263 #
14)	MA Endrin	6.065	6.734	72536265	372.8E6	67.722	52.980
16)	A 4,4'-DDD	0.000	6.854	0	2922343	N.D.	0.440 #
17)	MA 4,4'-DDT	6.430	7.153	133.4E6	718.1E6	136.874	108.189
20)	A Methoxychlor	6.977	7.610	162.2E6	832.0E6	297.239	242.410
21)	B Endrin ke...	0.000	7.761	0	2677324	N.D.	0.367 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD020921\
Data File : PD061143.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Feb 2021 11:41
Operator : DD\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampled :
PEM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 09 11:52:58 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012821.M
Quant Title : GC Extractables
QLast Update : Fri Jan 29 01:48:44 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

