

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD020924\
 Data File : PD081253.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Feb 2024 10:35
 Operator : AR\AJ
 Sample : P1379-01
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 RE137-CARBON-20240205

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 09 21:41:55 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012424.M
 Quant Title : GC Extractables
 QLast Update : Thu Jan 25 11:14:56 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.573	2.792	5510424	10538409	3.882	4.405
28)	SA Decachlor...	9.120	7.933	17790074	28206876	8.840	10.590

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD020924\
Data File : PD081253.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Feb 2024 10:35
Operator : AR\AJ
Sample : P1379-01
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
RE137-CARBON-20240205

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 09 21:41:55 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012424.M
Quant Title : GC Extractables
QLast Update : Thu Jan 25 11:14:56 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.573 min
Delta R.T.: 0.007 min
Response: 5510424
Conc: 3.88 ng/ml

Instrument :
ECD_D
ClientSampleId :
RE137-CARBON-20240205




#1 Tetrachloro-m-xylene

R.T.: 2.792 min
Delta R.T.: -0.004 min
Response: 10538409
Conc: 4.41 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.120 min
Delta R.T.: 0.007 min
Response: 17790074
Conc: 8.84 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.933 min
Delta R.T.: 0.000 min
Response: 28206876
Conc: 10.59 ng/ml