

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD021022\
 Data File : PD067997.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Feb 2022 09:58
 Operator : AR\AJ
 Sample : N1323-12
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEST-GPC2-BLANK-SPIKE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 11 06:01:48 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD020922CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Feb 09 15:28:35 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds						
3) MA	gamma-BHC...	4.152	4.660	58844339	248.7E6	20.667 20.554m
4) MA	Heptachlor	4.444	5.160	58106017	195.5E6	19.930 19.386m
5) MB	Aldrin	4.689	5.465	59040084	180.5E6	20.446 20.257
13) MA	Dieldrin	5.708	6.470	106.9E6	294.3E6	41.216 40.876
14) MA	Endrin	5.965	6.689	88847725	246.6E6	41.132 40.159
17) MA	4,4'-DDT	6.328	7.113	75460741	246.4E6	40.375 41.153

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD021022\
Data File : PD067997.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Feb 2022 09:58
Operator : AR\AJ
Sample : N1323-12
Misc :
ALS Vial : 5 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 11 06:01:48 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD020922CLP.M
Quant Title : GC Extractables
QLast Update : Wed Feb 09 15:28:35 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

