

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD021220\
Data File : PD057224.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Feb 2020 12:07
Operator : AJ\MA
Sample : L1422-10
Misc :
ALS Vial : 15 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
C5AX3

Manual Integrations
APPROVED

mohammad
2/13/2020 2:51:47 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 12 23:38:12 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021020CLP.M
Quant Title : GC Extractables
QLast Update : Tue Feb 11 04:58:40 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(13) Dieldrin (MA)
5.519min 3.118 ng/ml
response 3282254

(13) Dieldrin #2 (MA)
0.000min 0.000 ng/ml
response 0

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD021220\
Data File : PD057224.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Feb 2020 12:07
Operator : AJ\MA
Sample : L1422-10
Misc :
ALS Vial : 15 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
C5AX3

Manual Integrations
APPROVED

mohammad
2/13/2020 2:51:47 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 12 23:38:12 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021020CLP.M
Quant Title : GC Extractables
QLast Update : Tue Feb 11 04:58:40 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



QEdit

(13) Dieldrin (MA)
5.519min 3.118 ng/ml
response 3282254

(13) Dieldrin #2 (MA)
6.432min 1.809 ng/ml m
response 4865834

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD021220\
 Data File : PD057224.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Feb 2020 12:07
 Operator : AJ\MA
 Sample : L1422-10
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 C5AX3

Manual Integrations
 APPROVED

mohammad
 2/13/2020 2:51:47 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 12 23:38:12 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021020CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Feb 11 04:58:40 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.288	3.959	14787345	41549025	17.801	22.296 #
27) SA Decachlor...	7.958	8.996	28084922	60404784	27.072	26.721
Target Compounds						
13) MA Dieldrin	5.519	6.432	3282254	4865834	3.118	1.809m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

SS
 2/20/20