

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD021220\
 Data File : PD057217.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Feb 2020 09:50
 Operator : AJ\MA
 Sample : PB126681BS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PLCS81

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 12 23:35:39 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021020CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Feb 11 04:58:40 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.286	3.960	18843358	47789811	22.684	25.645
27)	SA Decachlor...	7.957	8.996	39416749	85175586	37.994	37.678
Target Compounds							
3)	MA gamma-BHC...	3.987	4.629	4815107	15204646	3.797	5.366 #
8)	B Heptachlo...	4.947	5.822	5153984	13630359	4.971	5.529
10)	B trans-Chl...	5.170	6.056	5450741	14509480	5.043	5.307
12)	B 4,4'-DDE	5.394	6.285	10927444	26094147	9.883	10.196
13)	MA Dieldrin	5.518	6.434	10425051	27418810	9.905	10.194
14)	MA Endrin	5.771	6.651	9055358	21442512	10.304	10.190
19)	B Endosulfa...	6.423	7.205	4603975	12799343	5.163	5.455

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD021220\
Data File : PD057217.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Feb 2020 09:50
Operator : AJ\MA
Sample : PB126681BS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 12 23:35:39 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021020CLP.M
Quant Title : GC Extractables
QLast Update : Tue Feb 11 04:58:40 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

