

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD021424\
 Data File : PD081345.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Feb 2024 17:16
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 14 01:35:58 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012424.M
 Quant Title : GC Extractables
 QLast Update : Thu Jan 25 11:14:56 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.565	2.794	72466639	133.3E6	51.048	55.719
2)	SA Decachlor...	9.117	7.934	97383291	136.0E6	48.390	51.042
Target Compounds							
2)	A alpha-BHC	4.023	3.296	127.1E6	216.4E6	51.172	56.385
3)	MA gamma-BHC...	4.356	3.625	125.0E6	209.1E6	51.054	55.912
4)	MA Heptachlor	4.949	3.966	121.6E6	193.4E6	49.290	56.079
5)	MB Aldrin	5.292	4.247	123.0E6	200.7E6	50.099	55.103
6)	B beta-BHC	4.551	3.920	50624440	84457273	50.032	54.049
7)	B delta-BHC	4.800	4.151	126.9E6	212.9E6	50.570	56.735
8)	B Heptachlo...	5.720	4.748	115.6E6	179.4E6	51.609	53.567
9)	A Endosulfan I	6.107	5.118	108.5E6	164.6E6	52.324	53.756
10)	B gamma-Chl...	5.976	4.998	114.8E6	177.6E6	52.449	54.286
11)	B alpha-Chl...	6.056	5.062	114.6E6	175.4E6	52.619	53.650
12)	B 4,4'-DDE	6.227	5.251	109.0E6	168.9E6	53.551	55.135
13)	MA Dieldrin	6.382	5.383	116.1E6	184.3E6	52.547	54.200
14)	MA Endrin	6.613	5.658	97759056	157.3E6	52.340	53.073
15)	B Endosulfa...	6.831	5.951	103.1E6	157.3E6	53.237	54.707
16)	A 4,4'-DDD	6.744	5.805	90469515	144.6E6	55.059	55.086
17)	MA 4,4'-DDT	7.060	6.056	94025419	142.5E6	52.797	53.723
18)	B Endrin al...	6.961	6.130	79970023	121.0E6	52.572	53.660
19)	B Endosulfa...	7.196	6.352	100.8E6	150.7E6	53.076	54.030
20)	A Methoxychlor	7.535	6.630	50573870	78509096	49.135	50.157
21)	B Endrin ke...	7.683	6.859	110.5E6	178.6E6	52.858	55.331
22)	Mirex	8.160	7.042	82309436	130.1E6	49.903	51.618

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD021424\
Data File : PD081345.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Feb 2024 17:16
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 14 01:35:58 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012424.M
Quant Title : GC Extractables
QLast Update : Thu Jan 25 11:14:56 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

