

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD021424\
Data File : PD081342.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Feb 2024 15:58
Operator : AR\AJ
Sample : PB158965BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB158965BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 14 01:35:11 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012424.M
Quant Title : GC Extractables
QLast Update : Thu Jan 25 11:14:56 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.565	2.794	25564840	44780283	18.009	18.720
28)	SA Decachlor...	9.116	7.934	34876208	46951126	17.330	17.627

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD021424\
Data File : PD081342.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Feb 2024 15:58
Operator : AR\AJ
Sample : PB158965BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB158965BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 14 01:35:11 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021424.M
Quant Title : GC Extractables
QLast Update : Thu Jan 25 11:14:56 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.565 min
Delta R.T.: 0.000 min
Response: 25564840
Conc: 18.01 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB158965BL




#1 Tetrachloro-m-xylene

R.T.: 2.794 min
Delta R.T.: -0.002 min
Response: 44780283
Conc: 18.72 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.116 min
Delta R.T.: 0.003 min
Response: 34876208
Conc: 17.33 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.934 min
Delta R.T.: 0.001 min
Response: 46951126
Conc: 17.63 ng/ml