

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD021523\
 Data File : PD073969.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Feb 2023 18:39
 Operator : AR\AJ
 Sample : PB150874BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 15 20:23:03 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012523CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Jan 25 02:23:06 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.555	2.776	42945744	22718795	18.135	17.898
27)	SA Decachlor...	9.103	7.932	105.1E6	45653049	38.107	37.932
Target Compounds							
2)	A alpha-BHC	3.964f	0.000	156668	0	<MDL	N.D. #
4)	MA Heptachlor	4.952	0.000	1503942	0	0.413	N.D. #
6)	B beta-BHC	0.000	3.869f	0	274645	N.D.	0.336
7)	B delta-BHC	4.782	0.000	2455875	0	0.752	N.D. #
8)	B Heptachlo...	5.656f	4.776f	328708	-49343	0.101	N.D. #
10)	B trans-Chl...	6.005f	0.000	10106667	0	3.139	N.D. #
11)	B cis-Chlor...	6.005f	0.000	10106667	0	3.032	N.D. #
12)	B 4,4'-DDE	6.189	0.000	187366	0	<MDL	N.D. #
20)	A Methoxychlor	7.464f	0.000	173038	0	0.116	N.D. #
21)	B Endrin ke...	7.703	0.000	20447	0	<MDL	N.D. #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD021523\
Data File : PD073969.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Feb 2023 18:39
Operator : AR\AJ
Sample : PB150874BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 15 20:23:03 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD012523CLP.M
Quant Title : GC Extractables
QLast Update : Wed Jan 25 02:23:06 2023
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.555 min
Delta R.T.: 0.000 min
Response: 42945744
Conc: 18.13 ng/ml




#1 Tetrachloro-m-xylene

R.T.: 2.776 min
Delta R.T.: 0.000 min
Response: 22718795
Conc: 17.90 ng/ml




#4 Heptachlor

R.T.: 4.952 min
Delta R.T.: 0.016 min
Response: 1503942
Conc: 0.41 ng/ml




#4 Heptachlor

R.T.: 0.000 min
Exp R.T.: 3.952 min
Response: 0
Conc: N.D.




#6 beta-BHC

R.T.: 0.000 min
Exp R.T. : 4.537 min
Response: 0
Conc: N.D.




#6 beta-BHC

R.T.: 3.869 min
Delta R.T.: -0.036 min
Response: 274645
Conc: 0.34 ng/ml




#7 delta-BHC

R.T.: 4.782 min
Delta R.T.: -0.004 min
Response: 2455875
Conc: 0.75 ng/ml




#7 delta-BHC

R.T.: 0.000 min
Exp R.T. : 4.137 min
Response: 0
Conc: N.D.




#8 Heptachlor epoxide

R.T.: 5.656 min
Delta R.T.: -0.049 min
Response: 328708
Conc: 0.10 ng/ml




#8 Heptachlor epoxide

R.T.: 4.776 min
Delta R.T.: 0.039 min
Response: -49343
Conc: N.D.




#10 trans-Chlordane

R.T.: 6.005 min
Delta R.T.: 0.042 min
Response: 10106667
Conc: 3.14 ng/ml




#10 trans-Chlordane

R.T.: 0.000 min
Exp R.T.: 4.988 min
Response: 0
Conc: N.D.




#11 cis-Chlordane

R.T.: 6.005 min
Delta R.T.: -0.038 min
Response: 10106667
Conc: 3.03 ng/ml




#11 cis-Chlordane

R.T.: 0.000 min
Exp R.T. : 5.052 min
Response: 0
Conc: N.D.




#20 Methoxychlor

R.T.: 7.464 min
Delta R.T.: -0.058 min
Response: 173038
Conc: 0.12 ng/ml




#20 Methoxychlor

R.T.: 0.000 min
Exp R.T. : 6.625 min
Response: 0
Conc: N.D.



#27 Decachlorobiphenyl

R.T.: 9.103 min
Delta R.T.: 0.004 min
Response: 105134956
Conc: 38.11 ng/ml



#27 Decachlorobiphenyl

R.T.: 7.932 min
Delta R.T.: 0.000 min
Response: 45653049
Conc: 37.93 ng/ml