

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD021622\
 Data File : PD068049.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Feb 2022 12:39
 Operator : AR\AJ
 Sample : PB142711BS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PLCS711

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 02/17/2022
 Supervised By :Ankita Jodhani 02/17/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 17 03:30:37 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021522CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Feb 16 09:16:50 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.451	4.151	26397887	211.0E6	20.106	22.838
27)	SA Decachlor...	8.218	9.330	52984550	255.2E6	53.331	52.629
Target Compounds							
2)	A alpha-BHC	3.891	4.546	7731634	66062548	4.328	4.965
3)	MA gamma-BHC...	4.175	4.837	7573315	67908362	4.504	5.584
4)	MA Heptachlor	4.475	5.362	7593146	62797923	5.016	5.891
5)	MB Aldrin	4.723	5.674	8127323	67378371	4.986	5.869
6)	B beta-BHC	4.425	4.999	4362480	31079582	5.788	5.975
7)	B delta-BHC	4.629	5.222	2954029	28438853	1.824	2.325m#
8)	B Heptachlo...	5.167	6.062	7625664	59883521	5.247	6.105
9)	A Endosulfan I	5.507	6.423	6793829	48059180	4.961	5.611
10)	B trans-Chl...	5.395	6.297	7732335	56977535	5.294	5.995m
11)	B cis-Chlor...	5.453	6.373	7411257	51846492	5.217	5.765
12)	B 4,4'-DDE	5.619	6.523	13458073	102.3E6	9.951	11.996
13)	MA Dieldrin	5.751	6.682	14509938	104.2E6	9.535	11.470
14)	MA Endrin	6.007	6.900	13024706	75736968	10.816m	11.475m
15)	B Endosulfa...	6.281	7.104	12115107	78294088	9.923m	11.183m
16)	A 4,4'-DDD	6.137	7.013	11616079	86867332	9.766	12.607 #
17)	MA 4,4'-DDT	6.377	7.316	8525770	55928144	9.240m	10.686
18)	B Endrin al...	6.453	7.230	10269858	62907858	11.766	12.551
19)	B Endosulfa...	6.665	7.454	11046633	57237070	9.969	9.340m
20)	A Methoxychlor	6.930	7.769	28753033	144.8E6	53.036	53.998
21)	B Endrin ke...	7.158	7.929	12779760	74974562	11.379	12.370

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD021622\
Data File : PD068049.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Feb 2022 12:39
Operator : AR\AJ
Sample : PB142711BS
Misc :
ALS Vial : 9 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 17 03:30:37 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021522CLP.M
Quant Title : GC Extractables
QLast Update : Wed Feb 16 09:16:50 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Reviewed By: Abdul Mirza 02/17/2022
Supervised By: Ankita Jodhani 02/17/2022

Manual Integrations
APPROVED

Instrument :
ECD_D
ClientSampled :
PLCS711