

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD021720\
 Data File : PD057341.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Feb 2020 12:23
 Operator : AJ\MA
 Sample : L1418-19
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 C5B21

Manual Integrations
 APPROVED

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 18 06:44:20 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021020CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Feb 11 04:58:40 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.284	3.958	7893784	18937414	9.503m	10.162m
27)	SA Decachlor...	7.954	9.018	11889845	7797665	11.461m	3.449m#
Target Compounds							
10)	B trans-Chl...	5.171	6.051	39360035	57027447	36.414m	20.859m#
11)	B cis-Chlor...	5.225	6.126	8732935	50448490	8.171m	19.692m#
12)	B 4,4'-DDE	5.393	6.281	71071658	130.2E6	64.277m	50.860m
16)	A 4,4'-DDD	5.903	6.772	273.0E6	664.8E6	394.055m	322.434
17)	MA 4,4'-DDT	6.138	7.068	11772946	20133042	18.092	10.604m#


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

mohammad

2/24/2020 2:50:49 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 18 06:44:20 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021020CLP.M
Quant Title : GC Extractables
QLast Update : Tue Feb 11 04:58:40 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Manual Integrations
APPROVED

```
Instrument :  
ECD_D  
ClientSampleId :  
C5B21
```