

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD021822\
 Data File : PD068090.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Feb 2022 14:01
 Operator : AR\AJ
 Sample : INDA201
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA2100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 18 15:00:20 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021822CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Feb 18 14:59:12 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.451	4.179	14664749	107.2E6	10.688	11.208
27)	SA Decachlor...	8.220	9.381	25328703	96750327	21.029	21.059
Target Compounds							
2)	A alpha-BHC	3.892	4.576	19429398	144.1E6	10.235	10.915
3)	MA gamma-BHC...	4.175	4.867	18729390	133.9E6	10.372	11.011
4)	MA Heptachlor	4.475	5.394	17238374	110.1E6	10.401	10.751
9)	A Endosulfan I	5.508	6.458	16777204	98617564	10.682	10.869
13)	MA Dieldrin	5.752	6.717	35362003	198.7E6	20.700	21.599
14)	MA Endrin	6.010	6.938	27300261	138.9E6	20.357	21.158
16)	A 4,4'-DDD	6.138	7.048	27906778	144.1E6	20.690	21.192
17)	MA 4,4'-DDT	6.379	7.351	22247420	100.9E6	19.616	20.191
20)	A Methoxychlor	6.932	7.804	65276524	229.4E6	103.232	102.869

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD021822\
Data File : PD068090.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Feb 2022 14:01
Operator : AR\AJ
Sample : INDA201
Misc :
ALS Vial : 11 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 18 15:00:20 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021822CLP.M
Quant Title : GC Extractables
QLast Update : Fri Feb 18 14:59:12 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

