

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD021825CLP.M
 Title : GC Extractables
 Last Update : Thu Feb 20 13:52:49 2025
 Response Via : Initial Calibration

Calibration Files

3 =PD087893.D 1 =PD087891.D 2 =PD087892.D
 4 =PD087894.D 5 =PD087895.D

Compound		3	1	2	4	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.156	2.223	2.187	2.131	2.112	2.162	E6	2.04
2) A	alpha-BHC						0.000		-1.00
3) MA	gamma-BHC (Lindane)						0.000		-1.00
4) MA	Heptachlor						0.000		-1.00
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I						0.000		-1.00
10) B	trans-Chlordane						0.000		-1.00
11) B	cis-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin						0.000		-1.00
14) MA	Endrin						0.000		-1.00
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD						0.000		-1.00
17) MA	4,4'-DDT						0.000		-1.00
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00
20) A	Methoxychlor						0.000		-1.00
21) B	Endrin ketone						0.000		-1.00
22)	Toxaphene-1	2.205	2.232	2.202	2.191	2.202	2.207	E4	0.69
23)	Toxaphene-2	3.018	3.056	3.077	3.127	3.159	3.087	E4	1.82
24)	Toxaphene-3	8.301	7.942	8.048	8.745	9.044	8.416	E4	5.56
25)	Toxaphene-4	6.097	5.808	5.848	6.418	6.723	6.179	E4	6.31
26)	Toxaphene-5	4.422	4.279	4.365	4.610	4.811	4.497	E4	4.75
27) SA	Decachlorobiphenyl	2.595	2.784	2.696	2.487	2.391	2.591	E6	6.08

Signal #2 Calibration Files

3 =PD087893.D 1 =PD087891.D 2 =PD087892.D
 4 =PD087894.D 5 =PD087895.D

Compound		3	1	2	4	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.381	1.515	1.460	1.331	1.258	1.389	E7	7.34
2) A	alpha-BHC						0.000		-1.00
3) MA	gamma-BHC (Lindane)						0.000		-1.00
4) MA	Heptachlor						0.000		-1.00
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I						0.000		-1.00
10) B	trans-Chlordane						0.000		-1.00
11) B	cis-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin						0.000		-1.00
14) MA	Endrin						0.000		-1.00
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD						0.000		-1.00
17) MA	4,4'-DDT						0.000		-1.00
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00
20) A	Methoxychlor						0.000		-1.00
21) B	Endrin ketone						0.000		-1.00
22)	Toxaphene-1	0.934	1.001	0.978	0.886	0.830	0.926	E5	7.50
23)	Toxaphene-2	0.989	1.121	1.071	0.951	0.901	1.006	E5	8.84

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
Method File : PD021825CLP.M
Title : GC Extractables
Last Update : Thu Feb 20 13:52:49 2025
Response Via : Initial Calibration

Calibration Files

3 =PD087893.D 1 =PD087891.D 2 =PD087892.D
4 =PD087894.D 5 =PD087895.D

	Compound	3	1	2	4	5	Avg		%RSD
24)	Toxaphene-3	3.217	3.198	3.223	3.186	3.113	3.187	E5	1.39
25)	Toxaphene-4	2.223	2.398	2.332	2.152	2.035	2.228	E5	6.46
26)	Toxaphene-5	1.653	1.556	1.631	1.665	1.660	1.633	E5	2.75
27) SA	Decachlorobiphenyl	1.191	1.333	1.260	1.124	1.063	1.194	E7	8.97

(#) = Out of Range