

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD022120\
 Data File : PD057421.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Feb 2020 17:29
 Operator : AJ\MA
 Sample : INDA401
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA401

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 24 16:07:32 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD022120CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Feb 24 16:06:26 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.227	3.896	34641192	81716967	40.565	40.236
27)	SA Decachlor...	7.865	8.895	71094376	172.0E6	78.348	78.504
Target Compounds							
2)	A alpha-BHC	3.649	4.280	51019789	129.5E6	42.909	42.140
3)	MA gamma-BHC...	3.921	4.560	46968078	130.8E6	41.090	40.653
4)	MA Heptachlor	4.201	5.061	34047304	131.8E6	40.667	40.984
9)	A Endosulfan I	5.195	6.097	38194191	114.6E6	41.039	40.993
13)	MA Dieldrin	5.434	6.352	87728727	253.5E6	84.262	82.973
14)	MA Endrin	5.685	6.568	70824880	201.1E6	82.479	82.808
16)	A 4,4'-DDD	5.821	6.694	63890914	202.1E6	83.519	82.126
17)	MA 4,4'-DDT	6.056	6.991	60589905	194.7E6	85.536	85.860
20)	A Methoxychlor	6.603	7.449	149.2E6	502.8E6	402.710	395.214

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD022120\
Data File : PD057421.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Feb 2020 17:29
Operator : AJ\MA
Sample : INDA401
Misc :
ALS Vial : 16 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 24 16:07:32 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD022120CLP.M
Quant Title : GC Extractables
QLast Update : Mon Feb 24 16:06:26 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

