

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD022120\
 Data File : PD057423.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Feb 2020 17:58
 Operator : AJ\MA
 Sample : INDA501
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA501

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 24 16:07:40 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD022120CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Feb 24 16:06:26 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.227	3.896	72045856	168.1E6	84.367	82.762
27)	SA Decachlor...	7.865	8.896	144.1E6	339.2E6	158.787	154.812
Target Compounds							
2)	A alpha-BHC	3.650	4.280	112.3E6	276.7E6	94.431	90.059
3)	MA gamma-BHC...	3.921	4.560	100.9E6	273.3E6	88.297	84.976
4)	MA Heptachlor	4.201	5.062	72180934	275.6E6	86.216	85.687
9)	A Endosulfan I	5.196	6.098	81155914	238.4E6	87.200	85.272
13)	MA Dieldrin	5.434	6.353	189.2E6	532.3E6	181.763	174.231
14)	MA Endrin	5.685	6.568	148.5E6	418.2E6	172.992	172.185
16)	A 4,4'-DDD	5.822	6.694	137.3E6	419.7E6	179.460	170.557
17)	MA 4,4'-DDT	6.055	6.991	134.1E6	414.8E6	189.347	182.958
20)	A Methoxychlor	6.603	7.450	304.2E6	986.4E6	821.016	775.367

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD022120\
Data File : PD057423.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Feb 2020 17:58
Operator : AJ\MA
Sample : INDA501
Misc :
ALS Vial : 18 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 24 16:07:40 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD022120CLP.M
Quant Title : GC Extractables
QLast Update : Mon Feb 24 16:06:26 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

